




Agilent 6400 Series Triple Quad LC/MS System

Installation Guide



Agilent Technologies

Notices

© Agilent Technologies, Inc. 2012

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

Manual Part Number

G3335-90150

Edition

Revision A, November 2012

Printed in USA

Agilent Technologies, Inc.
5301 Stevens Creek Blvd.
Santa Clara, CA 95051

Microsoft®, Windows® and Excel® are registered trademarks of Microsoft Corporation in the US and/or other countries.

Patents

This product is covered by U.S. patents 6,107,628 and 6,803,565 licensed from Battelle Memorial Institute.

Software Revision

This guide is valid for the B.06.00 revision or higher of the Agilent MassHunter Workstation Software - Data Acquisition program and compatible Qualitative Analysis and Quantitative Analysis programs, until superseded.

Note to Customer

This product may be used as a component of an in vitro diagnostic system if the system is registered with the appropriate authorities and complies with the relevant regulations. Otherwise, it is intended only for general laboratory use.

Warranty

The material contained in this document is provided “as is,” and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Restricted Rights Legend

U.S. Government Restricted Rights. Software and technical data rights granted to the federal government include only those rights customarily provided to end user customers. Agilent provides this customary commercial license in Software and technical data pursuant to FAR 12.211 (Technical Data) and 12.212 (Computer Software) and, for the Department of Defense, DFARS 252.227-7015 (Technical Data - Commercial Items) and DFARS 227.7202-3 (Rights in Commercial Computer Software or Computer Software Documentation).

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

In This Guide...

This guide contains information to install the Agilent 6400 Series Triple Quad LC/MS System.

1 Instrument Installation

This chapter describes the steps that are needed to install the Triple Quad instrument.

2 Software Installation and Configuration

This chapter describes the steps that are needed to install the MassHunter software.

3 Installation Completion and System Verification

In this chapter, you finish the hardware setup.

4 Installation and Verification of Other Sources

This chapter contains the steps needed to install and verify the operation of the APCI, APPI, MultiMode, and HPLC-Chip Cube sources.

5 Other Setup Tasks

This chapter contains tasks that are related to setting up your Triple Quad instrument.

6 Reference

This chapter contains parameters for verifying performance.

Content

1	Instrument Installation	7
	Step 1. Unpack and set up the Triple Quad instrument	8
	Step 2. Prepare to install the system	9
	Step 3a. Set up Agilent MS40+ pumps	13
	Step 3b. Set up the Edwards E2M18 pump	16
	Step 4. Connect the Triple Quad instrument	20
	Step 5. Install the spray chamber	26
	Step 6. Install the Agilent LC system	31
2	Software Installation and Configuration	33
	Step 1. Set up the computer system	34
	Step 2. Check PC network card configuration	35
	Step 3. Prepare for installation	37
	Step 4. Install the Data Acquisition program	38
	Step 5. Install the Qualitative Analysis program	42
	Step 6. Install the Quantitative Analysis program	42
	Step 7. Install and configure Microsoft Excel	43
	Step 8. Install Service Packs for Data Acquisition	43
	Step 9. Install Compliance programs (optional)	44
	Step 10. Install Quantitative Analysis Reporting	45
	Step 11. Check Excel Add-Ins for MassHunter	45
	Step 12. Configure the instrument	46
	Step 13. Copy the support folder to the computer	47
	Step 14. Confirm the LC firmware revision	49
	Step 15. Finish the installation	50
3	Installation Completion and System Verification	51
	Step 1. Condition the Agilent LC instrument	52
	Step 2. Prepare the performance evaluation samples	55
	Step 3. Review methods and run worklist	58

Step 4. Review methods and run worklist for bid specification (add-on service)	69
Step 5. Verify the Triple Quad sensitivity	72
To reset the Collision Cell gas flow rate	78
4 Installation and Verification of Other Sources	81
G1947B APCI Source	82
To install the G1947B APCI source	82
To verify the operation of the G1947B APCI source	83
G1978B Multimode Source	84
To install the G1978B multimode source	84
To verify the operation of the G1978B multimode source	86
G1971B APPI Source	87
To install the G1971B APPI source	87
To verify the operation of the G1971B APPI source	88
HPLC-Chip/MS Cube Interface	89
Step 1. Prepare performance verification sample	89
Step 2. Prepare the HPLC-Chip/MS	91
Step 3. Analyze the results	94
5 Other Setup Tasks	97
To start the Data Acquisition program	98
To stop MassHunter Engines	98
To stop MassHunter processes in case of unrecoverable errors	98
To view the log book	99
To configure the network interface card	100
To remove the MassHunter Workstation software	101
To reconfigure the instrument	102
6 Reference	105
6460/6430 Reserpine Checkout Method Acquisition Parameters	106
6490 Reserpine Checkout Method Acquisition Parameters	110

1

Instrument Installation

- Step 1. Unpack and set up the Triple Quad instrument 8
- Step 2. Prepare to install the system 9
- Step 3a. Set up Agilent MS40+ pumps 13
- Step 3b. Set up the Edwards E2M18 pump 16
- Step 4. Connect the Triple Quad instrument 20
- Step 5. Install the spray chamber 26
- Step 6. Install the Agilent LC system 31

This chapter describes the steps that are needed to install the Triple Quad instrument.



Figure 1 6490 Triple Quad LC/MS



Step 1. Unpack and set up the Triple Quad instrument

NOTE

Do not open the shipping containers until an Agilent representative is present to verify the contents of each carton. Warranty claims for missing items are honored only if an Agilent representative is on site to verify the contents of each shipping container as it is unpacked.

Checkout and tuning samples are shipped separately to comply with shipping regulations and to ensure prompt delivery. Make sure that these samples have been stored correctly, as indicated in the instructions in the box.

- 1** Carefully examine all containers for external signs of damage. If damage is discovered, immediately contact the carrier and the Agilent Service Management.

Retain shipping containers and material until contents are checked for completeness and instrument performance is verified.

- 2** Check off each item on the packing list, and verify the serial numbers.
- 3** Record the serial numbers in the installation documentation.
- 4** Report any discrepancies to Agilent Technologies.
- 5** Check that site preparation is complete, which includes power and gas supplies and chemical supplies needed for installation and performance verification.
- 6** Check the line voltage in the lab.

Verify the proper line voltage and outlet ratings according to the instructions given in the *Site Preparation Guide*.

Step 2. Prepare to install the system

Before you begin, check that you have these parts:

- Utility knife and needle nose pliers
- 1 Unpack and set up the Triple Quad instrument.
 - a Cut the tie wraps around the package. See [Figure 2](#).



Figure 2 Triple Quad instrument shipping container

- b Lift the boxes off the top of the instrument. See [Figure 3](#).

1 Instrument Installation

Step 2. Prepare to install the system




Figure 3 Instrument unboxed on shipping pallet (6460 pictured)

- c** Remove the front, top, right, back-left, and lower front-left cosmetic covers so that you can remove the shipping brackets. You do not need to remove the upper front-left cover.

To remove the top cover, remove the two screws at the back side of the top cover that secure the back of the top cover to the instrument chassis (see [Figure 4](#)). These screws are only needed when shipping the instrument. You do not need to reinstall the screws when you reinstall the top cover.



Figure 4 Remove shipping screws from back of top panel. The screws to remove are circled in this picture.

- d** Remove the shipping brackets (a total of 7). You will need to use a 13-mm wrench.




Figure 5 Triple Quad on shipping pallet with covers removed (6460 pictured)

1 Instrument Installation

Step 2. Prepare to install the system

- e Take an inventory of the contents of the packages.
 - f Record serial numbers. The serial number label for the Triple Quad instrument is located on the lower left corner to the left of the instrument power push button.
- 2 Use four people, one at each corner of the instrument, to lift the Triple Quad instrument onto a stable bench or table.

WARNING

Do not lift the Triple Quad without enough assistance to do so safely. The instrument weighs up to 115 kg (255 lbs). Be sure to use correct lifting procedures to avoid possible injury. Lift with your knees, not with your back. Keep your back straight while you bend your knees.

NOTE

Do NOT lift from under the power switch bracket on the front left corner of the instrument. The bracket will bend. See [Figure 6](#).



Figure 6 Do not lift from under power switch bracket


Step 3a. Set up Agilent MS40+ pumps

Do these steps if your instrument include one or more Agilent MS40+ pump.

Install two foreline pumps for 6490 and one foreline pump for 6420.

Before you begin, make sure that you have:

- Pump oil pan (p/n G1946-00034)
- Tygon 1/2-inch I.D. exhaust tubing (p/n G1946-80077)
- Agilent MS40+ foreline pump
- SW60 pump oil
- NW25 to 5/8-inch tube elbow (p/n G1960-20003)
- NW20/25 hinged clamp with wing nut (p/n 0100-1398)
- O-ring (p/n 0905-1592)
- pump power cord (p/n 8121-1753)



- 1 Unpack the Tygon 1/2-inch I.D. exhaust tubing (p/n G1946-80077) and Pump oil pan (p/n G1946-00034) from the shipping kit.
- 2 Remove the Agilent MS40+ foreline pump from its shipping container.

1 Instrument Installation

Step 3a. Set up Agilent MS40+ pumps

- 3 Set the Agilent MS40+ foreline pump into the Pump oil pan (p/n G1946-00034).

The pan is used to catch any small amounts of oil that can seep out of the pump. The pan can also hold all of the oil in the pump in case of a catastrophic seal failure.



- 4 Cut an appropriate length of the Tygon 1/2-inch I.D. exhaust tubing (p/n G1946-80077).
- 5 Find the NW25 to 5/8-inch tube elbow (p/n G1960-20003) and the O-ring (p/n 0905-1592). Put the O-ring on the exhaust adaptor.
- 6 Locate one NW20/25 hinged clamp with wing nut (p/n 0100-1398), open the swing clamp, and use it to clamp the exhaust adaptor onto the oil mist filter. Connect the Tygon 1/2-inch I.D. exhaust tubing (p/n G1946-80077) to the exhaust adaptor.
- 7 Add the entire 1-liter bottle of SW60 pump oil to the foreline pump.

CAUTION

The pump is shipped dry. Oil must be used for proper operation. Failure to add oil before you pump down the system will damage the foreline pump.

- 8 Plug the pump power cord (p/n 8121-1753) into the power inlet of the foreline pump. Secure the pump power cord in place using the lock mechanism.
- 9 For 6490: Plug the other end into the pump expander box port labeled Foreline. Secure the power cord in place using the lock mechanism. Do this for both pumps. See Figure 7.



Figure 7 Pump expander box ports.

Step 3b. Set up the Edwards E2M18 pump

Do these steps if your instrument includes an Edwards E2M18 pump.

Before you begin, check that you have:

- Edwards mist filter and instructions (p/n 3162-1056)
- Edwards oil drain kit and instructions (p/n 3162-1057)
- NW25 to 5/8-inch tube elbow (p/n G1960-20003)
- NW20/25 hinged clamp with wing nut (p/n 0100-1398)
- O-ring (p/n 0905-1592)
- metal foreline pump hose (p/n G1969- 20123)
- pump oil pan (p/n G1946-00034)
- Tygon 1/2-inch I.D. exhaust tubing (p/n G1946-80077)
- NW25 centering ring (p/n 0100-1551)

If the system is installed with a G3215A Mass Spec Bench, please follow the instructions to install the pumps inside the pump enclosure of the Mass Spec Bench.

- 1** Unpack the mist filter, exhaust hose, and pump oil drip from the shipping kit:
- 2** Remove the foreline pump from its shipping container:
 - a** Cut the straps that secure the cardboard box to the wooden pallet.
 - b** Lift off the cardboard box.
 - c** Set aside the two bottles of pump oil and funnel.
 - d** Set aside the oil return kit. The felt plugs and screen inside the packaging are not needed and may be discarded.
 - e** Use a pair of needle nose pliers to remove the metal clips that secure the pump to the wooden pallet.
- 3** Set the pump into the [pump oil pan \(p/n G1946-00034\)](#).

The pan is used to catch any small amounts of oil that may seep out of the pump. The pan can also contain all of the oil in the pump in case of a catastrophic seal failure.

4 Fill the foreline pump with Inland 45 fluid.

The foreline pump is drained of vacuum pump oil prior to shipment. Remove the oil fill cap and use the funnel to add Inland 45 pump fluid until the level in the site glass is approximately 2/3 full.

5 Install the oil mist filter.

Remove the plugged outlet connector and install the new KF-25 fitting that comes with the pump. Do not use the outlet nozzle as described in the Edwards manual and on the warning tag. Install the mist filter on the outlet port KF-25 fitting. See figure on the next page.

Make sure the mist filter drain plug is facing the foreline pump ballast valve (toward the motor end of the pump).

6 Install the oil return line:

- a** Remove the drain plug from the mist filter and install the drain adapter.
- b** Install the brass restrictor into the black oil return tubing.
- c** Connect the oil return tubing between the foreline pump ballast port fitting and the oil mist filter drain adapter. Make sure the end of the tubing with the brass restrictor is installed at the ballast port fitting.
- d** Use the plastic hose clips and a pair of needle nose pliers to secure the oil return tubing.

1 Instrument Installation

Step 3b. Set up the Edwards E2M18 pump




Figure 8 Example foreline pump. Note that your pump may look slightly different.

7 Check that the ballast valve is closed.

Closing the ballast valve prevents additional air from being introduced to the pump. This allows the foreline pump to reach a lower operating pressure.

CAUTION

The ballast valve must be opened (fully counterclockwise) periodically to allow oil trapped in the mist filter to return to the pump. If the oil is not returned to the pump, the pump will eventually run dry and be damaged.

Step 3b. Set up the Edwards E2M18 pump

- 8 Find the NW25 to 5/8-inch tube elbow (p/n G1960-20003) and the O-ring (p/n 0905-1592). Put the O-ring on the exhaust adaptor.
- 9 Locate one NW20/25 hinged clamp with wing nut (p/n 0100-1398), open the swing clamp, and use it to clamp the exhaust adaptor onto the oil mist filter. Connect the Tygon 1/2-inch I.D. exhaust tubing (p/n G1946-80077) to the exhaust adaptor.
- 10 Attach one end of the metal foreline pump hose (p/n G1969- 20123) using one NW20/25 hinged clamp with wing nut (p/n 0100-1398) and one NW25 centering ring (p/n 0100-1551) to the LC/MS mainframe.
- 11 Connect the other end of the metal foreline pump hose to the inlet fitting on the foreline pump using one NW20/25 hinged clamp with wing nut (p/n 0100-1398) and one NW25 centering ring (p/n 0100-1551).
- 12 Plug the power cord from the foreline pump into the power outlet labeled **Foreline** on the AC board.


Step 4. Connect the Triple Quad instrument

- 1 Connect the rough hose (p/n G1969-20123) that is supplied in the ship kit to the KF-25 foreline pump hose connection on the back left of the instrument.
- 2 Connect the foreline pump hose to the KF-25 inlet adapter on the foreline pump.
- 3 Use the supplied hook/loop fastener tape to secure the drain bottle into a secondary containment tub.
- 4 Connect the spray chamber drain tube to the 1-inch fitting on the drain bottle. The drain tube can also be routed out the front of the Triple Quad if needed.
- 5 Attach the calibrant delivery system (CDS)/inlet module waste tubing from the solvent selection valve on the Triple Quad to the ¼-inch fitting on the drain bottle.
- 6 Connect a length of the ½-inch Tygon tubing to the ½-inch fitting on the drain bottle, and then connect the other end of the tubing to a vent connection that is separate from the vent used for the foreline pump.

CAUTION


The drain bottle vent must be located away from the foreline pump vent to prevent the foreline pump exhaust from contaminating the Triple Quad spray chamber.

Step 4. Connect the Triple Quad instrument

**Figure 9** Drain bottle**7** Install the gas filter(s):

For Triple Quad instruments without AJS ESI source:

- a** Cut an appropriate length of the 1/4-inch PTFE gas tubing (G1946-80078) and connect to the nitrogen gas supply regulator or nitrogen generator.
- b** Use the supplied nut and ferrule kit (5183-0392) to attach the other length of tubing to the RMSN-4 gas filter. This configures the filter in series with the supply tubing. See [Figure 10](#) as an example for assembling the gas trap in the correct configuration.

**Figure 10** Gas filter connected in series

1 Instrument Installation

Step 4. Connect the Triple Quad instrument

For 6460 and 6490 with AJS ESI source:

- a** Cut an appropriate length of the 1/4-inch PTFE gas tubing (p/n G1946-80078) and connect to the nitrogen gas supply regulator or nitrogen generator.
- b** Locate the 2 1/4-inch brass T-unions (0100-0088) and loosen the brass nuts on the T-unions.
- c** Connect the other end of the supply tubing to the T-union.
- d** Cut four 4-inch lengths of PTFE tubing from the supplied 10 meter length (G1946-80078).
- e** Connect two of the 4 inch lengths to each side of the T-unions. Use the supplied nut and ferrule kit (5183-0392) to attach the other side 4 inch the tubing to the two RMSN-4 gas filters. This configures the filters in parallel. See [Figure 11](#) as an example for assembling the gas traps in the correct configuration.




Figure 11 Gas filter connected in parallel

Step 4. Connect the Triple Quad instrument




Figure 12 Power and gas connections on back left of Triple Quad

- 8 Set the pressure on the nitrogen supply regulator at 80 to 100 psi (550 to 690 kPa). Turn on the nitrogen gas for a few minutes to purge the tubing and gas conditioner before you connect the nitrogen gas supply to the Triple Quad. Regulator pressures above 110 psi (760 kPa) cause nitrogen waste due to release from the bleed valve on the flow-control module.
- 9 Use the tubing cutters (p/n 8710-1709) supplied in the ship kit to cut an appropriate length of the 1/8-inch copper supply tubing (p/n 5180-4196) to connect from the outlet fitting on the collision cell nitrogen tank regulator to the nitrogen gas conditioner (p/n RMSN-2) for the collision cell gas, and then cut a second length of the 1/8 -inch copper supply tubing to connect from the nitrogen gas conditioner to the collision cell gas inlet fitting on the

1 Instrument Installation

Step 4. Connect the Triple Quad instrument

left side of the Triple Quad (see Figure 12). Secure the gas filters in a vertical position with tie wraps.




Figure 13 Power and electrical connection at back left side of Triple Quad

- 10 Set the pressure on the Collision Cell supply regulator at 10 to 30 psi (70 to 210 kPa).

CAUTION

Do not attach the Collision Cell gas fitting to the nitrogen supply regulator at 80 to 100 psi (550 to 690 kPa). Pressure to the Collision Cell Gas Flow Module greater than 30 psi (210 kPa) can cause fewer collisions to be available for secondary collision induced dissociation (CID) and poor peak shapes, as well as poor high vacuum. The Collision Cell can use the same nitrogen supply as the regular supply (i.e. for drying gas, nebulizer, CDS pressure) if it is the required purity (99.999%) but it must be down-regulated to 30 psi (210 kPa).

- 11** Connect the remote start cable to its connector on the back left side of the Triple Quad. See [Figure 13](#).
- 12** Connect the foreline pump power cord into the foreline pump connector on the back left side of the Triple Quad (see [Figure 13](#)).
- 13** Verify the front power switch is in the OFF position.
- 14** Plug the Triple Quad power cord into the Triple Quad power connector on the Triple Quad and the other end of the Triple Quad power cord into the wall outlet. See [Figure 13](#).
- 15** For 6490 only: Plug the Power Expansion box power cord into the wall outlet.

Step 5. Install the spray chamber

Agilent Jet Stream ESI Source (6460/6490 only)

Before you begin, check that you have these parts:

- Agilent Jet Stream (AJS) ESI source
 - ESI-L Low Concentration Tuning Mix (p/n G1969-85000)
- 1 Remove the top cover by loosening the T10 lock screw and lifting off the instrument chassis.
 - 2 Remove the AJS ESI source from the packaging inside the instrument chassis next to the e-Module. See [Figure 14](#).




Figure 14 Packaging for AJS ESI source for 6460. For 6490, the AJS ESI is shipped outside of the instrument chassis.

- 3 Remove the foil that covers the spray chamber mount, and remove the shipping cover from the Electrospray spray chamber.
- 4 Put the nebulizer in the nebulizer adjustment fixture supplied in the shipping kit and check that the nebulizer needle is properly adjusted. The nebulizer needle should be even with the end of the nebulizer nozzle.
- 5 Install the nebulizer in the spray chamber.

- 6 Install the spray chamber on the spray chamber mount, close the spray chamber, and fasten the latch. You may need to adjust the latch to ensure that the O-ring seals completely. Use a ¼-inch x 5/16-inch wrench to loosen the lock nut, and then adjust the latch to the proper fit, and then tighten the lock nut so that the latch maintains its adjustment.
- 7 Connect the high voltage and vaporizer cables to the connectors on the left side panel of the instrument.
- 8 Connect the 1/8-inch nebulizing gas tubing from the Triple Quad mainframe to the nebulizer gas fitting.
- 9 Connect the 1/8-inch sheath gas tubing from the Triple Quad mainframe to the sheath gas fitting on the AJS ESI source.
- 10 Connect the PEEK tubing from the selection valve (inside the front cover) to the top of the nebulizer.




Figure 15 Connections for AJS ESI source

1 Instrument Installation

Step 5. Install the spray chamber

11 Rinse the calibrant bottle three times with hot (60°C) water and then three times with methanol. Then rinse it once with acetonitrile.

12 Add the ESI-L Low Concentration Tuning Mix to the calibrant bottle.

CAUTION

Never use aliphatic, aromatic or halogenated hydrocarbons in the CDS. These solvents are not compatible with the O-ring in the CDS.

13 Turn on the Triple Quad front power switch to start the pump-down of the Triple Quad.

The foreline pump(s) will become quieter within a few seconds.

Agilent G1948B Electrospray Source

Before you begin, check that you have these parts:

- Agilent G1948B Electrospray interface (the G1948A Electrospray interface is not supported)
- Nebulizer spacer kit (p/n G1946-60254)
- ESI-L Low Concentration Tuning Mix (p/n G1969-85000)

1 Remove the foil that covers the spray chamber mount, and remove the shipping cover from the Electrospray spray chamber.

2 Install the nebulizer spacer as follows:

- a** Use a flat-bladed screwdriver to remove the two nebulizer shoulder screws from the top of the spray chamber.
- b** Use the two m3x8 Torx T10 screws to install the nebulizer spacer onto the top of the spray chamber.
- c** Install the two nebulizer shoulder screws into the top of the nebulizer spacer.

3 Put the nebulizer in the nebulizer adjustment fixture supplied in the shipping kit and check that the nebulizer needle is properly adjusted. The nebulizer needle should be even with the end of the nebulizer nozzle.

4 Install the nebulizer in the spray chamber.

5 Install the spray chamber on the spray chamber mount, close the spray chamber, and fasten the latch. You may need to adjust the latch to ensure that the O-ring seals completely. Use a ¼-inch x 5/16-inch wrench to loosen

the lock nut, adjust the latch to the proper fit, and then tighten the lock nut so that the latch maintains its adjustment.

- 6 Connect the 1/8-inch nebulizing gas tubing from the Triple Quad mainframe to the nebulizer gas fitting.
- 7 Connect the PEEK tubing from the selection valve (inside the front cover) to the top of the nebulizer.
- 8 Rinse the calibrant bottle three times with hot (60°C) water and then three times with methanol. Then rinse it once with acetonitrile.
- 9 Add the Electrospray calibrant to the calibrant bottle.

CAUTION

Never use aliphatic, aromatic or halogenated hydrocarbons in the CDS. These solvents are not compatible with the O-ring in the CDS.

- 10 Turn on the Triple Quad front power switch to start the pump-down of the Triple Quad.

The foreline pump(s) will become quieter within a few seconds.

G1978B Multimode Source

- 1 Install these parts from the Multimode and APCI enablement kits that ship with the G1978B Multimode source into the Aux module on the Triple Quad LC/MS:
 - APCI High Voltage Power Supply (p/n G1946-80058)
 - Multimode HV board (p/n G1960-61015)
 - Valve board—APCI HV PS cable (p/n G1960-60802)
 - Valve board—APCI Needle Interlock cable (p/n G1960-60856)
 - Multimode HV board cable (p/n G1960-60858)
 - Multimode Power Data cable (p/n G1960-60878)
 - MMI-L Low Concentration Tuning Mix (p/n G1969-85020)
- 2 Connect the multimode and corona needle cables to the respective connectors on the left side panel of the Triple Quad LC/MS.
- 3 Put the nebulizer in the nebulizer adjustment fixture supplied in the shipping kit and check that the nebulizer needle is properly adjusted.

The nebulizer needle should be even with the end of the nebulizer nozzle.

1 Instrument Installation

Step 5. Install the spray chamber

- 4 Install the nebulizer in the spray chamber.
- 5 Install the spray chamber on the spray chamber mount, close the spray chamber, and fasten the latch. You may need to adjust the latch to ensure that the O-ring seals completely. Use a ¼-inch x 5/16-inch wrench to loosen the lock nut, adjust the latch to the proper fit, and then tighten the lock nut so that the latch maintains its adjustment.
- 6 Connect the 1/8-inch nebulizing gas tubing from the Triple Quad mainframe to the nebulizer gas fitting.
- 7 Connect the PEEK tubing from the selection valve (inside the front cover) to the top of the nebulizer.
- 8 Rinse the calibrant bottle three times with hot (60°C) water and then three times with methanol. Then rinse it once with acetonitrile.
- 9 Add the Multimode calibrant to the calibrant bottle.

CAUTION

Never use aliphatic, aromatic or halogenated hydrocarbons in the CDS. These solvents are not compatible with the O-ring in the CDS.

- 10 Turn on the Triple Quad front power switch to start the pump-down of the Triple Quad.

The foreline pump(s) will become quieter within a few seconds.

Step 6. Install the Agilent LC system

Before you begin, check that you have these parts:

- HPLC grade methanol, or better
 - HPLC grade water, or better
- 1** Install the Agilent LC system. Follow its installation documents. Other LCs are not supported.
 - 2** Connect the remote start cable from the Triple Quad to a Remote Start connector on the LC sampler, pump or other detector module only.
 - 3** Install an Agilent G1369 LAN Interface card into the LC detector module if required. If a detector module is not present, install the Agilent G1369 LAN Interface card into the pump module. Before you install the LAN Interface card, set switches 5 and 6 to the ON position. This will force the Agilent LAN Interface Card to use its default IP address and subnet mask.


NOTE

Note: For a G1315C/D DAD SL, the LAN interface is built into the module main board. To set the G1315C/D to the proper IP address, set switch positions 7 and 8 to the ON or UP position. By default, switch positions 7 and 8 are in the down position. Refer to the *Agilent 1200 Series Diode Array and Multiple Wavelength Detector User Manual* (part number G1315-90012) for more information.

The default IP address is 192.168.254.11 and the default subnet mask is 255.255.255.0.


1 Instrument Installation

Step 6. Install the Agilent LC system



CAUTION

Do not use any other LAN interface card besides the G1369 (e.g. the older Jet Direct card J4100A).



2 Software Installation and Configuration

- Step 1. Set up the computer system 34
- Step 2. Check PC network card configuration 35
- Step 3. Prepare for installation 37
- Step 4. Install the Data Acquisition program 38
- Step 5. Install the Qualitative Analysis program 42
- Step 6. Install the Quantitative Analysis program 42
- Step 7. Install and configure Microsoft Excel 43
- Step 8. Install Service Packs for Data Acquisition 43
- Step 9. Install Compliance programs (optional) 44
- Step 10. Install Quantitative Analysis Reporting 45
- Step 11. Check Excel Add-Ins for MassHunter 45
- Step 12. Configure the instrument 46
- Step 13. Copy the support folder to the computer 47
- Step 14. Confirm the LC firmware revision 49
- Step 15. Finish the installation 50

If you purchased a bundled Triple Quad instrument system, the MassHunter Workstation software has been installed for you. In this case, skip to “[Step 7. Install and configure Microsoft Excel](#)” and complete the steps that remain.



Step 1. Set up the computer system

- 1 Set up the PC:
 - a Unpack the PC and put it on a suitable benchtop or table.
 - b Connect the keyboard, mouse, monitor and printer to the computer.
- 2 Verify that two network cards are available in the PC.

One of these is usually a “built-in” card, and the second is located in an accessory slot. Follow the installation documentation supplied with the PC to install and verify components as needed.
- 3 Install the network hub:
 - a Put the hub on the table next to the PC.
 - b Plug the power supply into the hub.
 - c Connect the power supply to a power outlet.
- 4 Connect one end of a Category 5, shielded twisted pair (STP) cable to any port (1 to 7) on the network hub. Connect the other end of the cable to the 10/100 Base-T network card in slot 1 of the PC. The card should be labeled **LC/MS**.

Do not connect to port 8 on the hub/switch.
- 5 Connect a second Category 5 STP LAN cable from one of the open ports 1 to 7 on the hub to the Agilent LAN Card in the pump or detector.
- 6 Connect a third Category 5 STP LAN cable from one of the open ports 1 to 8 on the hub to the LAN interface connection on the LAN/MS control card in the Triple Quad instrument.
- 7 Install the printer. See the installation documentation supplied with the printer.
- 8 Turn on the printer, hub, monitor, and PC, in that order.

The computer comes installed with software and the appropriate drivers for the installed accessories.
- 9 When the computer starts, create a user account. This account will be used to acquire data. Do not use the Admin account for this purpose.

Step 2. Check PC network card configuration

- 1 Click **Start > Control Panel**, and then under **Network and Internet**, click **View network status and tasks**.

If Control Panel displays in Small icons view, click **Network and Sharing Centers**.

- 2 Click **Change adapter settings**.
- 3 From the list of network connections, double-click the network connection that corresponds to the instrument LAN connection.
- 4 Scroll down the list to **Internet IP Protocol 4 (TCP/IPv4)** (or **Internet IP Protocol**), select it, and then click **Properties**.




Figure 16 LCMS Connection Properties dialog box - Networking tab

2 Software Installation and Configuration

Step 2. Check PC network card configuration

- 5 Make sure the Network Adapter is set to these settings:

IP address: 192.168.254.1

Subnet mask: 255.255.255.0




Figure 17 Internet Protocol Version 4 (TCP/IP) Properties dialog box

Step 3. Prepare for installation

- 1 Use the **Uninstall or change a programs** command in Control Panel to remove programs that you will be upgrading. *Do not remove programs for which you need only to install a service pack.*
 - Excel, if you need to upgrade (only Excel 2010 or Office 2010 is supported)
 - Agilent MassHunter Workstation Qualitative Analysis
 - Agilent MassHunter Workstation Quantitative Analysis
 - Agilent MassHunter Workstation Quantitative Analysis Reporting
 - Agilent MassHunter Workstation Quantitative Analysis Compliance
- 2 Back up the data, method and worklist files in the **D:\MassHunter** folder. Use the back-up method of your choice.

The uninstallation program keeps your data, method and worklist files. It also backs up your tuning and hardware setting files, which are restored during installation. Do not move, remove or rename the D:\MassHunter folder, or your tuning and hardware settings will not be saved.

- 3 Run Windows Update to make sure you have the latest critical updates and security fixes.

Make sure Windows Update is completed before you continue.

CAUTION

If you do not run Windows Update, your system may be vulnerable to security problems. Confirm that the LAN power management has not been activated by a Windows Update program.

Step 4. Install the Data Acquisition program

The Data Acquisition program is pre-installed on the shipping PC when ordered together with the instrument as a bundle. If the program is already installed, continue at “[Step 7. Install and configure Microsoft Excel](#)” on page 43.

If the upgrade to the Data Acquisition program is for a service pack only, continue at “[Step 5. Install the Qualitative Analysis program](#)” on page 42.

The installation of the Data Acquisition program includes installation of Optimizer and Study Manager. Do not separately install these programs.

- 1 Check that no other program is running on your system, including Windows Update.
- 2 Put the Data Acquisition installation disc into the disc drive.

The welcome screen appears.

- 3 If this is an upgrade installation, click **Uninstall MassHunter Data Acquisition and Optimizer Software**, then follow the instructions on the screen to remove the current program.
- 4 Click **Install MassHunter Data Acquisition Software**, then follow the instructions on the screen to install the software.
 - In the License Agreement screen, click **I accept the terms of the license agreement**, then click **Next**.
 - In the Choose Destination Location screen, click **Next** to accept the default destination location.

The default is **C:\Program Files (x86)\Agilent\MassHunter\Workstation\Acq.**

- In the Choose location for customer data files screen, check that the default of D:\MassHunter is displayed in the text box, then click **Next**. See [Figure 18](#).




Figure 18 Choose location for customer data files screen

- In the Start Copying Files screen, review the settings then click **Next**.

The Setup Status screen now appears.

- If you get a message that indicates that the screen resolution is *not* set to the required resolution, click **OK**.
- If you see a Notepad window that prompts you to install other options, close the Notepad window.
- When the InstallShield Wizard Complete message is displayed, click **Finish**.

5 Click **Install Microsoft .NET 4.0** and follow the instructions, if any, to install Microsoft .NET Framework 4.0.


6 If this installation is a software-only upgrade:

- a** Click **Upgrade MS Firmware**.
- b** Press **Enter** to accept the default IP address for the Triple Quad instrument (192.168.254.12).

2 Software Installation and Configuration

Step 4. Install the Data Acquisition program


- c When prompted, type **Y** for yes to proceed with update.



```
D:\WWWDemoCD (B15)\QQQFW_47\QqQUpdate.exe
D:\WWWDemoCD (B15)\QQQFW_47\QqQUpdate.exe version 1.00
Enter instrument ip: 192.168.254.12
192.168.254.12 communications check PASSED
SmartcardIO detected.

MSD controller @ 192.168.254.12 will be updated
!!>>DO NOT DISRUPT POWER TO MSD DURING THE UPDATE.<<!!
OK to proceed (Y=YES): _
```

- 7 If the instrument is using an AJS ESI and this is a new installation:
- a In Windows Explorer, find the folder **\Support\QQQFirmware\QQQFW_0x_xx**, where *x_xx* represents the version of the shipping firmware shipping, then run **QqQUpdate.exe**.
 - b Type the IP address of the Triple Quad instrument (192.168.254.12) and press **Enter**.
 - c When prompted, enter the country-specific nominal input voltage. Verify with a Digital Multi-Meter (DM) using the AC Voltage measuring function.



```
C:\Documents and Settings\Administrator\My Documents\QQQ_B2018\QQQFW_05_06\QqQUpdate.exe
Enter instrument ip: 192.168.254.12
192.168.254.12 communications check PASSED
SmartcardIO detected.

MSD controller @ 192.168.254.12 will be updated
!!>>DO NOT DISRUPT POWER TO MSD DURING THE UPDATE.<<!!
OK to proceed (Y=YES): y
Clearing old firmware.
Retreiving 192.168.254.12:QQQDefaultParms.000

Nominal line voltage:
0: (Shipping Default)
1: 190 volts
2: 200 volts
3: 208 volts
4: 220 volts
5: 230 volts
6: 240 volts
7: 250 volts
The current setting is #0, 0 volts
Line Voltage is 0, this is only for shipment.
press [return] to use current one or
Input the item number to select line voltage:_
```

For example, in the United States, the closest nominal voltage for the Triple Quad instrument is 208V so you would select option 3.

CAUTION

The AJS ESI uses 95% input voltage duty cycle to heat properly. Improper settings can severely hamper its performance.

- d** When prompted, confirm input voltage set in the previous step.

```

C:\Documents and Settings\Administrator\My Documents\QQQ_B2018\QQQFW_05_06\QQQ...
!>>>DO NOT DISRUPT POWER TO MSD DURING THE UPDATE.<C:\Documents and Settings\Administrator\My Docume
OK to proceed (Y=YES): y
Clearing old firmware.
Retreiving 192.168.254.12:QQQDefaultParms.000

Nominal line voltage:
0: <Shipping Default>
1: 190 volts
2: 200 volts
3: 208 volts
4: 220 volts
5: 230 volts
6: 240 volts
7: 250 volts
The current setting is #3, 208 volts
press (return) to use current one or
Input the item number to select line voltage:3
The line voltage is set to 208.
Is this correct? (y/n):y
Sending 192.168.254.12:/home/sc4/qqq/fileupdate
Sending 192.168.254.12:qqq/qqqServer
Sending 192.168.254.12:qqq/qqq_DSP_FW.ldr
update the files and reboot qqqServer.
  
```

- e** Reboot the LC/MS interface card, or turn on and off the instrument, when prompted.

Once the firmware update has completed, you are prompted to close the DOS window.

- 8** *Optional.* Click **Install MassHunter Data Acquisition Manuals** to copy the user guides to the location of your choice.

The installation program automatically copies the MassHunter user guides onto your computer so that they can be opened from online Help. The user guides are also available on the installation disc in the Manuals folder. Click this button only if you want an additional set of user guides copied to the location of your choice.

Step 5. Install the Qualitative Analysis program

- Refer to the *Offline Qualitative and Quantitative Analysis Installation Guide* to install the Qualitative Analysis program.

CAUTION

You must use the **Run as administrator** option to install the program. Failure to do so can result in an incomplete installation, which can cause the system to run unreliably.

Do not install any other MassHunter program until instructed to do so.

Step 6. Install the Quantitative Analysis program

- Refer to the *Offline Qualitative and Quantitative Analysis Installation Guide* to install the Quantitative Analysis program.

CAUTION

You must use the **Run as administrator** option to install the program. Failure to do so can result in an incomplete installation, which can cause the system to run unreliably.

Do not install any other MassHunter program until instructed to do so.

Step 7. Install and configure Microsoft Excel

If you are doing an upgrade installation, continue at “Step 8. Install Service Packs for Data Acquisition” on page 43.

The MassHunter Data Acquisition program supports:

- Excel 2010
- Office 2010
- Refer to the *Offline Qualitative and Quantitative Analysis Installation Guide* to install and configure Excel.

Do not install any other MassHunter program until instructed to do so.

Step 8. Install Service Packs for Data Acquisition

- 1 Put the Data Acquisition installation disc into the disc drive.

If you received a separate Service Pack installation disc, insert that disc instead.

- 2 If the folder **Service Packs** exists on the installation disc, then:
 - a Open the **Service Packs** folder. If the folder contains another folder, such as **SP1**, open that folder.
 - b Locate the installation file, which ends with **SP x.exe**.
 - c Right-click the installation file and select **Run as administrator**.
 - d Follow the instructions to install the service pack.

Step 9. Install Compliance programs (optional)

Do these steps if you need to enable compliance. You must install the MassHunter Data Acquisition program before you install the MassHunter Quantitative Analysis Compliance program.

1 Install the Quantitative Compliance program:

Refer to the *Offline Qualitative and Quantitative Analysis Installation Guide* to install the Quantitative Compliance program.

2 Enable Compliance for the Data Acquisition program.

Refer to the *MassHunter Data Acquisition Compliance Software Quick Start Guide* for details.

The compliance programs are installed and enabled.

Step 10. Install Quantitative Analysis Reporting

- 1 Refer to the *Offline Qualitative and Quantitative Analysis Installation Guide* to install the Qualitative Analysis Reporting program.

CAUTION

You must use the **Run as administrator** option to install the program. Failure to do so can result in an incomplete installation, which can cause the system to run unreliably.

Do not install any other MassHunter program until instructed to do so.

- 2 If you plan to use a PCL6 printer, download the print driver patch from <http://support.microsoft.com/kb/935843>.

This patch is needed to correctly print reports in A4 format to HP PCL6 printers.

- 3 Set the Microsoft Image Writer as the default printer:

- a Click **Start > Devices and Printers**.
- b Right-click **Microsoft Office Document Image Printer** and click **Set as Default Printer**.

The Microsoft Image Writer formats the Excel workbooks three to four times faster for networked printers because the formatting page set-up information is stored on the local computer. The report is processed by the Microsoft Image Writer, but it is actually printed on the printer that is selected in the MassHunter program.

CAUTION

If you do not set up a default printer, such as the Microsoft Office Document Image Printer, reporting can fail.

Step 11. Check Excel Add-Ins for MassHunter

- Refer to the *Offline Qualitative and Quantitative Analysis Installation Guide* to verify the installation of Excel Add-Ins for MassHunter.

Step 12. Configure the instrument

In this step, you configure the Triple Quad instrument and the HPLC instrument.

- 1 Open the **Agilent MassHunter** folder on the desktop, and then double-click **Acq Tools**.
- 2 Double-click the **Instrument Configuration** tool.
- 3 In the Instrument Configuration dialog box:
 - a If you want to change the name of the instrument, type a new name for **Instrument Name**.
 - b Mark the **Mass Spectrometer** check box.
 - c Click **Agilent 6400 Series Triple Quadrupole**.
 - d Mark the **Agilent LC System Access** check box.
 - e Click **Device Config** and configure the LC system.
 - f Click **OK**.




Figure 19 Instrument Configuration Dialog Box

- 4 When configuration is completed, click **OK**.

Step 13. Copy the support folder to the computer

In this step, you copy the **\Support** folder from the Data Acquisition installation disk to the **D:\MassHunter** folder.

- 1 In a Windows Explorer window, open the installation disk drive.
- 2 Right-click the **Support** folder and click **Copy**.
- 3 Right-click the **D:\MassHunter** folder and click **Paste**.
- 4 In Windows Explorer, right-click the **D:\MassHunter\Support** folder and click **Properties**.




Figure 20 Support Properties dialog box

- 5 Clear the **Read-only** check box, and then click **Apply**.
The Confirm Attribute Changes dialog box is displayed.

2 Software Installation and Configuration

Step 13. Copy the support folder to the computer




Figure 21 Confirm Attribute Changes dialog box

- 6** Click **Apply changes to this folder, subfolders and files**, and then click **OK**.

Step 14. Confirm the LC firmware revision

- 1 Start the Data Acquisition program.
- 2 Click **File > Print > Instrument Configuration Report**.
- 3 Click **Screen**.
- 4 Click **OK**.
- 5 Go to
http://www.chem.agilent.com/_layouts/agilent/downloadFirmware.aspx?wid=69761 and check that you have the most recent firmware for your instrument.
- 6 If you need to update your LC firmware, follow the instructions on the web site. Make sure you use the latest firmware update tool, as instructed on the web site.
- 7 Check that the firmware for the G1369A Agilent LAN Interface card is the latest version available:
 - a Start a Command Prompt in Windows.
 - b Type `telnet 192.168.254.11` in the Command Prompt window and press **Enter**.
 - c Type `/` in the Telnet window.


The LAN Interface card displays its current configuration information.

The LAN Interface card requires 1.10 firmware or higher. If an upgrade is needed, please contact your Agilent representative.

Step 15. Finish the installation

- 1 Create a system recovery disc of your installed system.
Follow the instructions for the Backup Solution Bundle (G1030-64002) included in the Triple Quad LC/MS System bundle.
- 2 Install or enable any other program that you will use with your MassHunter workstation. Refer to the installation guide or quick start guide for these programs.
- 3 Make sure that the trust settings in Excel are set for all MassHunter users, including the guest account that is set up for compliance. See the *Offline Qualitative and Quantitative Analysis Installation Guide* for more information.
- 4 Install the programs that are part of the Agilent Software Backup Solution and do a system back-up of the computer.

Refer to the *H1172A Backup Solution Installation and User Guide* and the *Agilent Software Backup Solution Hardware and Software Installation Checklist* for details.



3 Installation Completion and System Verification

- Step 1. Condition the Agilent LC instrument 52
- Step 2. Prepare the performance evaluation samples 55
- Step 3. Review methods and run worklist 58
- Step 4. Review methods and run worklist for bid specification (add-on service) 69
- Step 5. Verify the Triple Quad sensitivity 72
- To reset the Collision Cell gas flow rate 78

This chapter contains the steps needed to complete the hardware setup and verify the performance specifications of the installed system.



Step 1. Condition the Agilent LC instrument

This step is required if you have a 6490. It is optional if you purchased the G6061A Performance Specification Service.

Before you begin, check that you have:

- 18 MΩ, HPLC grade deionized water (p/n 8500-2236)
- high-purity HPLC grade isopropanol
- premixed LC flushing solvent (50:25:15:10 isopropanol/acetonitrile/cyclohexane/dichloromethane) (p/n G1969-85026)
- high-purity HPLC grade acetonitrile (p/n G2453-85050)
- high-purity HPLC grade methanol (p/n 8500-1867)
- high purity formic acid (p/n G2453-85060)

1 Flush the LC system.

This flushing procedure is very effective at reducing background contaminants from the LC systems when they are interfaced with a mass spectrometer. The flushing procedure must be done for each solvent channel to ensure contaminants are removed from separate degasser and pumping channels.

- 2** Remove any column(s) and bypass the flow cell for the UV detector, unless it is rated for high-pressure (400 bar) operation. Route the outlet capillary to a waste container. Don't try to re-use solvents used for the flushing procedure.
- 3** Use a handheld controller or the Data Acquisition program to set up the pump module to deliver deionized water for 15 minutes at 3 mL/minute to remove any salt buffers. Do 5 injections with deionized water to clean the injector path and sample loop.
- 4** Set up the pump module to deliver HPLC grade isopropanol for 5 minutes at 3 mL/minute. Do 5 injections with isopropanol to clean the injector path and sample loop.
- 5** Set up the pump module to deliver the premixed LC flushing solvent (50:25:15:10 isopropanol/acetonitrile/cyclohexane/dichloromethane) for 15 minutes at 3 mL/minute. Do 5 injections with the flushing solvent to clean the injector path and sample loop.

Alternatively, while the Triple Quad instrument pumps down, the flushing solvent can be pumped at a low flow rate (e.g. 0.1 mL/minute) overnight.

CAUTION

Make sure that the flushing solvent does not stay in the LC for an extended period of time (for example, several days) because the cyclohexane may be harmful to the pump seals.

- 6 Set up the pump module again to deliver HPLC grade isopropanol for 5 minutes at 3 mL/minute. Do 5 injections with isopropanol to clean the injector path and sample loop.
- 7 Install the column used for installation verification (p/n 827700-902) and wet the column with 100% methanol.

All columns shipped with the Triple Quad instrument are shipped dry. Upon installation, these columns must be primed or wetted with Methanol to properly condition the column.

- 8 Set up the pump module again to deliver HPLC grade methanol for at least 30 minutes at 0.5 mL/minute. Make sure that the solvent stream selector valve on the Triple Quad instrument is set to waste position or the outlet tubing from the column is *not* connected to the instrument.

If the HPLC Stack has a Thermostatted Column Compartment (TCC), set the temperature of the column compartment to 60°C to reduce the back pressure and to help remove any contaminants on the column.

- 9 Prepare mobile phase.

NOTE

Up to 0.1 percent formic acid can be added to the solvent that you just prepared to help ionize the reserpine checkout sample. Do not add the formic acid until after you prepare the Reserpine checkout sample.

Prepare 1 liter of 70:30 acetonitrile:water with 0.1 percent formic acid. Make sure the flow rate is set to 0.4 mL/minute to condition the column with checkout mobile phase.

- 10 Do 5 injections with the blank premixed mobile phase to clean the injector path and sample loop.

3 Installation Completion and System Verification

Step 1. Condition the Agilent LC instrument

NOTE

Use solvents that are at a minimum HPLC grade. Solvents that are acceptable for most LC applications may contain high levels of background that are detectable by the more sensitive Triple Quad instrument. LC solvents used with the Triple Quad instrument should be rated for both HPLC and pesticide, environmental, or GC/MS analyses. Use the highest purity solvents you can obtain. Acceptability of solvents must be empirically determined.

Step 2. Prepare the performance evaluation samples

In this step, you dilute the supplied performance evaluation samples to the concentrations needed for the Triple Quad system checkout. Use the diluted samples within a day of dilution. Refrigerate the intermediate (first) dilution in the supplied bottle.

- Always rinse the graduated pipettes and volumetric flasks thoroughly with deionized water before and between each use.
- Use polypropylene labware for preparing performance evaluation samples, since glass vessels introduce unacceptable levels of sodium.
- Always rinse the autosampler vials and caps with the solvent mix used for sample dilution before filling them with the performance verification samples. This minimizes any background that can be contributed by the vials and caps. The vials may be run uncapped if the septa are found to be a source of background contamination.

Required Parts

- 1 mL graduated pipette (p/n 9301-1423)
 - 2 × 100 mL volumetric flasks (p/n 9301-1344)
 - ES/APCI Positive Ion Performance Standard, Agilent G2423A
 - Plastic bottle for storing first dilution (p/n 9301-1433)
- 1 Transfer 1 mL of 5 ng/μL reserpine (p/n G2423A) to a 100 mL volumetric flask. Use a clean graduated pipette.
 - 2 Dilute to the 100 mL mark with 30:70 water:acetonitrile. Transfer 1 mL of the first dilution to a 100 mL volumetric flask. Use a clean graduated pipette.
 - 3 Dilute to the 100 mL mark with 30:70 water:acetonitrile.

This provides 500 fg/μL at 2 μL injection volume, which equals the final 1 pg/μL reserpine concentration that is needed for performance verification.

Table 1 6490 Performance Verification Summary, MRM Mode

6490 with AJS ESI Positive MRM Mode	
Sample	Reserpine, 5 ng/μL
Concentration after dilution	500 fg/μL
Injection volume	2 μL

3 Installation Completion and System Verification

Step 2. Prepare the performance evaluation samples

Table 1 6490 Performance Verification Summary, MRM Mode (continued)

6490 with AJS ESI Positive MRM Mode	
Total sample amount injected	1 pg
Sample order number (p/n)	G2423A
Solvent	30:70 water:acetonitrile
Method	6490 ESI_AJT Pos MRM Checkout.m
Worklist	6490 ESI_AJT Pos MRM Checkout.wkl
Performance specifications	10,000:1 (1x RMS)

Table 2 6430 and 6460 Performance Verification Summary, MRM Mode

	6460 with AJS ESI Positive MRM Mode	6430, 6460 ES G1948B Positive MRM Mode
Sample	Reserpine, 5 ng/μL	Reserpine, 5 ng/μL
Concentration after dilution	500 fg/μL	500 fg/μL
Injection volume	2 μL	2 μL
Total sample amount injected	1 pg	1 pg
Sample order number (p/n)	G2423A	G2423A
Solvent	30:70 water:acetonitrile	30:70 water:acetonitrile
Method	6460 ESI_AJT Pos MRM Checkout.m	6400 ESI Pos MRM Checkout.m
Worklist	6460 ESI_AJT Pos MRM Checkout.wkl	6400 ESI Pos MRM Checkout.wkl
Performance specifications	1000:1 (1x RMS)	300:1 (1 x RMS)

Table 3 6410/6420 Performance Verification Summary, MRM Mode

	ES G1948B Positive MRM Mode	ES G1978B Positive MRM Mode
Sample	Reserpine, 5 ng/μL	Reserpine, 5 ng/μL
Concentration after dilution	500 fg/μL	500 fg/μL
Injection volume	2 μL	2 μL
Total sample amount injected	1 pg	1 pg
Sample order number (p/n)	G2423A	G2423A
Solvent	30:70 water:acetonitrile	30:70 water:acetonitrile
Method	6400 ESI Pos MRM Checkout.m	MMI-ES Pos MRM Reserpine Checkout.m
Worklist	6400 ESI Pos MRM Checkout.wkl	MMI-ES Pos Reserpine Checkout.wkl
Performance specifications	150:1 (1 x RMS)	150:1 (1 x RMS)

Step 3. Review methods and run worklist

Before you begin, copy the methods and worklists according to your instrument model:

Table 4 Methods and worklists to copy

If you have...	Copy...	From...	To...
6490 with AJS ESI	6490 ESI_AJT Pos MS2 Scan.m 6490 ESI_AJT Pos MRM Checkout.m	D:\MassHunter\Support\ Checkout\6490 Methods and Worklist\Pos	D:\MassHunter\Methods
	6490 ESI_AJT Pos MRM Checkout.wkl	D:\MassHunter\Support\ Checkout\6490 Methods and Worklist\Pos	D:\MassHunter\Worklists
6460 with AJS ESI	6460 ESI_AJT Pos MS2 Scan.m 6460 ESI_AJT Pos MRM Checkout.m	D:\MassHunter\Support\ Checkout\6460 Methods and Worklist\Pos	D:\MassHunter\Methods
	6460 ESI_AJT Pos MRM Checkout.wkl	D:\MassHunter\Support\ Checkout\6460 Methods and Worklist\Pos	D:\MassHunter\Worklists
6410, 6420, 6430, 6460 ESI	6400 ESI Pos MRM Reserpine Checkout.m 6400 ESI Pos MS2 Scan.m	D:\MassHunter\Support\ Checkout\6400 Methods and Worklist\Pos\ESI Pos MRM Checkout	D:\MassHunter\Methods
	ESI Pos Reserpine Checkout.wkl	D:\MassHunter\Support\ Checkout\6400 Methods and Worklist\Pos\ESI Pos MRM Checkout	D:\MassHunter\Worklists

Depending on your instrument model, go to:

- “For 6490 - AJS ESI, Positive MRM Mode” on page 60
- “For 6460 - AJS ESI, Positive MRM Mode” on page 63
- “For 6410/6420/6430/6460 - G1948B Electrospray Interface, Positive MRM Mode” on page 65
- “For 6410/6420 - G1978B Multimode Interface, Positive MRM Mode” on page 67.

3 Installation Completion and System Verification

Step 3. Review methods and run worklist

For 6490 - AJS ESI, Positive MRM Mode

- 1 Verify that the Collision Cell gas flow rate produces a high vacuum gauge reading in the range of 5.0 to 5.8×10^{-5} torr.

To view the high vacuum gauge reading, start the Data Acquisition program, change the **Context** to **Tune**, and then click the **Manual Tune > Cell** tab. Check that the high vacuum gauge reading is within range. If not, see [“To reset the Collision Cell gas flow rate”](#) on page 78.

- 2 Start an **Autotune**.

After the autotune has completed, you may need to wait up to 30 minutes to allow for the calibrant solution to be pumped out of the Triple Quad. This minimizes any background signal attributable to the calibrant. In addition, you can sonicate the nebulizer in a small graduated cylinder filled with acetonitrile for 10 minutes.

NOTE

Parameters in the method can be modified by your Agilent representative for optimal response.

- 3 Change the Context to **Acquisition**, click **File > Load** and then load the method **6490 ESI_AJT Pos MRM Checkout.m**. The checkout method includes the following acquisition parameters. Edit if needed.

Table 5 6490 ESI_AJT Pos MRM Checkout.m parameters

Parameter	Value										
Injection volume	2 μ L										
Autosampler temperature	4°C										
Needle wash	flushport (25:75 water:methanol, 0.1% formic acid), 10 seconds										
Mobile Phase	A = water, 0.1% formic acid B = acetonitrile, 0.1% formic acid										
Gradient	<table><tr><th>Time</th><th>%B</th></tr><tr><td>0.2</td><td>10</td></tr><tr><td>0.8</td><td>98</td></tr><tr><td>1.0</td><td>98</td></tr><tr><td>1.1</td><td>98</td></tr></table>	Time	%B	0.2	10	0.8	98	1.0	98	1.1	98
Time	%B										
0.2	10										
0.8	98										
1.0	98										
1.1	98										
Run time	2 minutes										

Table 5 6490 ESI_AJT Pos MRM Checkout.m parameters (continued)

Parameter	Value
Flow rate	0.4 mL/minute
Drying gas temperature	200°C (can be optimized, typically between 130°C to 200°C)
Drying gas flow	12 L/minute (can be optimized, typically between 12 to 15 L/minute)
Nebulizer pressure	20 psi (can be optimized, typically between 20 psi to 25 psi)
Sheath gas temperature	400°C
Sheath gas flow	12 L/minute
Capillary voltage	3500 V
Nozzle voltage (AJS ESI)	500 V (can be optimized, typically between 0 to 500 V)
Collision cell energy	41 V (can be optimized, typically between 37 to 42 V)
Fragmentor voltage	fixed; determined by iFunnel and autotune
Accelerator voltage	3 V cell (can be optimized, typically between 1 to 3 V)
Delta EMV	200 V (can be optimized, typical between 200 to 350 V)
MS1 Resolution	Wide
MS2 Resolution	Unit

- 4 Make sure that for channels A and B, 90:10 water with 5 mM formic acid:acetonitrile is selected as the LC solvent.
- 5 Click **Method > Open** and then load the method **6490 ESI_AJT Pos MS2 Scan.m**.
- 6 Put the vials into the LC autosampler.
 - Position #1: An empty, uncapped vial
 - Position #2: A vial containing the solvent used for dilution (this is the solvent blank)
 - Position #3: A vial containing the reserpine sample (500 fg/μL)
- 7 Click **Worklist > Open**, and then load the worklist: **6490 ESI_AJT Pos MRM Checkout.wkl**.

3 Installation Completion and System Verification

Step 3. Review methods and run worklist

The worklist is set up to do one injection of the solvent blank using the **6490 ESI_AJT Pos MS2 Scan.m** method in order to collect background ion data, and then using the **6490 ESI_AJT Pos MRM Checkout.m** method for the remaining runs, one injection of the empty vial, five injections of the solvent blank, and five injections of the reserpine sample. After that, it runs the **6490 ESI_AJT Pos MRM Checkout.m** method to run five more injections of the reserpine sample with 0.1 µL injection volume.

For 6460 - AJS ESI, Positive MRM Mode

- 1 Verify that the Collision Cell gas flow rate produces a high vacuum gauge reading in the range of 1.9 to 2.3×10^{-5} torr.

To view the high vacuum gauge reading, start the Data Acquisition program, change the **Context** to **Tune**, and then click the **Manual Tune > Cell** tab. Check that the high vacuum gauge reading is within range. If not, see [“To reset the Collision Cell gas flow rate”](#) on page 78.

- 2 Start an **Autotune**.

After the autotune has completed, you may need to wait up to 30 minutes to allow for the calibrant solution to be pumped out of the Triple Quad. This minimizes any background signal attributable to the calibrant. In addition, you can sonicate the nebulizer in a small graduated cylinder filled with acetonitrile for 10 minutes.

- 3 Change the Context to **Acquisition**, click **Method > Open** and then load the method **6460 ESI_AJT Pos MRM Checkout.m**. The checkout method includes the following acquisition parameters:

Table 6 6460 ESI_AJT Pos MRM Checkout.m parameters

Parameter	Value
Injection volume	2 μ L
Run time	1 minute
Flow rate	isocratic from channel A and B 30:70, at 0.4 mL/minute
Drying gas temperature	325°C
Drying gas flow	10 L/minute
Nebulizer pressure	20 psi
Sheath gas temperature	400°C
Sheath gas flow	12 L/minute
Capillary voltage	4000 V
Nozzle voltage	500 V
Collision cell energy	41 V (can be optimized, typically between 37 to 41 V)
Fragmentor voltage	150 V (can be optimized, typically between 140 to 250 V)

3 Installation Completion and System Verification

Step 3. Review methods and run worklist

Table 6 6460 ESI_AJT Pos MRM Checkout.m parameters (continued)

Parameter	Value
Delta EMV	200 V (can be optimized, typical between 200 to 350 V)
MS1 Resolution	Wide
MS2 Resolution	Unit

NOTE

Parameters in the method can be modified by your Agilent representative for optimal response.

4 Click **Method > Open** and then load the method **6460 ESI_AJT Pos MS2 Scan.m**.

5 Put the vials into the LC autosampler.

- Position #1: An empty, uncapped vial
- Position #2: A vial containing the solvent used for dilution (this is the solvent blank)
- Position #3: A vial containing the reserpine sample (500 fg/μL)

6 Click **Worklist > Open**, and then open the worklist: **6460 ESI_AJT Pos MRM Checkout.wkl**.

The worklist is set up to do one injection of the solvent blank using the **6460 ESI_AJT Pos MS2 Scan.m** method in order to collect background ion data, and then using the **6460 ESI_AJT Pos MRM Checkout.m** method for the remaining runs, one injection of the empty vial, five injections of the solvent blank, and five injections of the reserpine sample.

For 6410/6420/6430/6460 - G1948B Electrospray Interface, Positive MRM Mode

- 1 Verify that the Collision Cell gas flow rate produces a high vacuum gauge reading in the range of 1.9 to 2.3×10^{-5} torr.

To view the high vacuum gauge reading, start the Data Acquisition program, change the **Context** to **Tune**, and then click the **Manual Tune > Cell** tab. Check that the high vacuum gauge reading is within range. If not, see [“To reset the Collision Cell gas flow rate”](#) on page 78.

- 2 Start an **Autotune**.

After the autotune has completed, you may need to wait up to 30 minutes to allow for the calibrant solution to be pumped out of the Triple Quad. This minimizes any background signal attributable to the calibrant. In addition, you can sonicate the nebulizer in a small graduated cylinder filled with acetonitrile for 10 minutes.

- 3 Change the **Context** to **Acquisition**, click **Method > Open** and then load the method **6400 ESI Pos MRM Checkout.m**. The checkout method includes the following acquisition parameters:

Table 7 6400 ESI Pos MRM Checkout.m parameters

Parameter	Value
Injection volume	2 μ L
Run time	1 minute
Flow rate	isocratic from channel A and B 30:70, at 0.4 mL/minute
Drying gas temperature	325°C
Drying gas flow	10 L/minute (can be optimized, typically between 10 to 13 L/minute)
Nebulizer pressure	35 psi
Capillary voltage	4000 V
Collision cell energy	37 V (can be optimized, typically between 37 to 41 V)
Fragmentor voltage	240 V (can be optimized, typically between 140 to 250 V)
Delta EMV	200 V (can be optimized, typical between 200 to 350 V)
MS1 Resolution	Wide
MS2 Resolution	Unit

3 Installation Completion and System Verification

Step 3. Review methods and run worklist

NOTE

Parameters in the method can be modified by your Agilent representative for optimal response.

- 4 Edit the method to ensure that for channels A and B, 30:70 water:acetonitrile with 5mM formic acid is selected as the LC solvent.
- 5 Click **Method > Open** and then load the method **6400 ESI Pos MS2 Scan.m**.
- 6 Put the vials into the LC autosampler.
 - Position #1: An empty, uncapped vial
 - Position #2: A vial containing the solvent used for dilution (this is the solvent blank)
 - Position #3: A vial containing the reserpine sample (500 fg/μL)
- 7 Click **Worklist > Open**, and then load the worklist: **6400 ESI Pos MRM Checkout.wkl**.

The worklist is set up to do one injection of the solvent blank using the **6400 ESI Pos MS2 Scan.m** method in order to collect background ion data, and then using the **6400 ESI Pos MRM Checkout.m** method for the remaining runs, one injection of the empty vial, five injections of the solvent blank, and five injections of the reserpine sample.

For 6410/6420 - G1978B Multimode Interface, Positive MRM Mode

- 1 Verify that the Collision Cell gas flow rate produces a high vacuum gauge reading in the range of 2.7 to 3.3×10^{-5} torr.

To view the high vacuum gauge reading, start the Data Acquisition program, change the **Context** to **Tune**, and then click the **Cell** tab. Check that the high vacuum gauge reading is within range. If not, see [“To reset the Collision Cell gas flow rate”](#) on page 78.

- 2 Start an **Autotune**.

After the autotune has completed, you may need to wait up to 30 minutes to allow for the calibrant solution to be pumped out of the Triple Quad. This minimizes any background signal attributable to the calibrant. In addition, you can sonicate the nebulizer in a small graduated cylinder filled with acetonitrile for 10 minutes.

- 3 Change the Context to **Acquisition**, click **Method > Open** and then load the method **MMI-ES Pos MRM Reserpine Checkout.m**. The checkout method includes the following acquisition parameters:

Table 8 MMI-ES Pos MRM Reserpine Checkout.m parameters

Parameter	Value
Injection volume	2 μ L
Run time	1 minute
Flow rate	isocratic from channel A, at 0.4 mL/minute
Drying gas temperature	250°C
Drying gas flow	10 L/minute (can be optimized, typically between 10 to 13 L/minute)
Nebulizer pressure	60 psi
Capillary voltage	2000 V
Collision cell energy	37 V (can be optimized, typically between 37 to 41 V)
Fragmentor voltage	240 V (can be optimized, typically between 140 to 250 V)
Delta EMV	400 V (can be optimized, typical between 200 to 400 V)
MS1 Resolution	Wide
MS2 Resolution	Unit

3 Installation Completion and System Verification

Step 3. Review methods and run worklist

- 4 Edit the method to ensure that for channel A, the 75:25 methanol:water solution with 5mM ammonium formate is selected as the LC solvent.
- 5 Click **Method > Open** and then load the method **MMI-ES Pos MS2 Scan.m**.
- 6 Edit the method to ensure that for channel A, 75:25 methanol:water solution with 5mM ammonium formate is selected as the LC solvent.
- 7 Put the vials into the LC autosampler.
 - Position #1: An empty, uncapped vial
 - Position #2: A vial containing the solvent used for dilution (this is the solvent blank)
 - Position #3: A vial containing the reserpine sample (500 fg/μL)
- 8 Click **Worklist > Open**, and then load the worklist: **MMI-ES Pos Reserpine Checkout.wkl**.

The worklist is set up to do one injection of the solvent blank using the **ESI Pos MS2 Scan.m** method in order to collect background ion data, and then using the **ESI Pos MRM Reserpine Checkout.m** method for the remaining runs, one injection of the empty vial, five injections of the solvent blank, and five injections of the reserpine sample.

Step 4. Review methods and run worklist for bid specification (add-on service)

The G6061A Performance Specification Service is available only at the time of original Agilent 6400 Series Triple Quad LC/MS System bundle purchase.

The steps in this topic are done by an Agilent representative if the G6061A service was purchased.

For 6490 with G6061A - AJS ESI, Positive MRM Mode, do the steps in [“For 6490 - AJS ESI, Positive MRM Mode”](#) on page 60.

For 6420, 6430, or 6460 with G6061A - ESI, Positive MRM Mode (AJS ESI), do the steps in this topic.

- 1 Verify that the Collision Cell gas flow rate produces a high vacuum gauge reading in the range of 1.9 to 2.3×10^{-5} torr for two turbo systems (6430 or 6460) and 2.7 to 3.3×10^{-5} torr for single turbo system (6420).

To view the high vacuum gauge reading, start the Data Acquisition program, change the **Context** to **Tune**, and then click the **Manual Tune > Cell** tab. Check that the high vacuum gauge reading is within range. If not, see [“To reset the Collision Cell gas flow rate”](#) on page 78.

- 2 Start an **Autotune**.

After the autotune has completed, you may need to wait up to 30 minutes to allow for the calibrant solution to be pumped out of the Triple Quad. This minimizes any background signal attributable to the calibrant. In addition, you can sonicate the nebulizer in a small graduated cylinder filled with acetonitrile for 10 minutes.

- 3 Change the **Context** to **Acquisition**, click **Method > Open**, and then load the method **6061 ESI Pos MRM Checkout.m**. The checkout method includes the following acquisition parameters. Edit if needed.

Table 9 6061 ESI Pos MRM Checkout.m parameters

Parameter	Value
Injection volume	2 μ L
Autosampler temperature	4°C
Needle wash	flushport (25:75 water:methanol, 0.1% formic acid), 10 seconds

3 Installation Completion and System Verification

Step 4. Review methods and run worklist for bid specification (add-on service)

Table 9 6061 ESI Pos MRM Checkout.m parameters (continued)

Parameter	Value												
Mobile Phase	A = water, 0.1% formic acid B = acetonitrile, 0.1% formic acid												
Gradient (for 1290)	<table> <tr> <th>Time</th><th>%B</th></tr> <tr> <td>0.2</td><td>10</td></tr> <tr> <td>0.8</td><td>98</td></tr> <tr> <td>1.0</td><td>98</td></tr> <tr> <td>1.1</td><td>98</td></tr> <tr> <td colspan="2">Stop time: 2 minutes</td></tr> </table>	Time	%B	0.2	10	0.8	98	1.0	98	1.1	98	Stop time: 2 minutes	
Time	%B												
0.2	10												
0.8	98												
1.0	98												
1.1	98												
Stop time: 2 minutes													
Gradient (for 1260)	<table> <tr> <th>Time</th><th>%B</th></tr> <tr> <td>0.2</td><td>10</td></tr> <tr> <td>1.0</td><td>98</td></tr> <tr> <td>2.0</td><td>98</td></tr> <tr> <td>2.1</td><td>98</td></tr> </table>	Time	%B	0.2	10	1.0	98	2.0	98	2.1	98		
Time	%B												
0.2	10												
1.0	98												
2.0	98												
2.1	98												
Run time	4 minutes												
Flow rate	0.4 mL/minute												
Drying gas temperature	350°C (can be optimized, typically between 325°C to 350°C)												
Drying gas flow	10 L/minute (can be optimized, typically between 10 to 12 L/minute)												
Nebulizer pressure	35 psi (can be optimized, typically between 20 psi to 35 psi)												
Sheath gas temperature (AJS ESI)	400°C												
Sheath gas flow	12 L/minute (AJS ESI)												
Capillary voltage	4000 V												
Nozzle voltage (AJS ESI)	500 V (can be optimized, typically between 0 to 500 V)												
Collision cell energy	41 V (can be optimized, typically between 37 to 42 V)												
Fragmentor voltage	150 V (can be optimized, typical between 150 V to 240 V)												
Accelerator voltage	4 V cell (can be optimized, typically between 2 to 5 V)												
Delta EMV	200 V (can be optimized, typical between 200 to 350 V)												

Step 4. Review methods and run worklist for bid specification (add-on service)

Table 9 6061 ESI Pos MRM Checkout.m parameters (continued)

Parameter	Value
MS1 Resolution	Wide
MS2 Resolution	Unit

NOTE

Parameters in the method can be changed by your Agilent representative for optimal response.

- 4 Make sure that for channels A and B, 90:10 water with 5 mM formic acid:acetonitrile are selected as the LC solvent.
- 5 Click **Method > Open**, and then load the method **6061 ESI Pos MS2 Scan.m**.
- 6 Put the vials into the LC autosampler.
 - Position #1: An empty, uncapped vial
 - Position #2: A vial containing the solvent used for dilution (this is the solvent blank)
 - Position #3: A vial containing the reserpine sample (500 fg/μL)
- 7 Click **Worklist > Open**, and then load the worklist: **6061 ESI Pos MRM Checkout.wkl**.

The worklist is set up to do one injection of the solvent blank using the **6061 ESI Pos MS2 Scan.m** method in order to collect background ion data, and then using the **6061 ESI Pos MRM Checkout.m** method for the remaining runs, one injection of the empty vial, five injections of the solvent blank, and five injections of the reserpine sample. After that, it runs the **6061 ESI Pos MRM Checkout.m** method to run five more injections of the reserpine sample with 0.1 μL injection volume.

Step 5. Verify the Triple Quad sensitivity

	Sample Name	Sample Position	Data File	Sample Type	Level Name	Inj Vol (µl)	Comment
1	☐ Background Scan	Vial 2	D:\MassHunter\Data\Checkout\6460\BackgroundScan001.d	Blank		1	
2	Air Blank	Vial 1	D:\MassHunter\Data\Checkout\6460\AirBlank001.d	Blank		1	
3	Solvent Blank	Vial 2	D:\MassHunter\Data\Checkout\6460\SolventBlank001.d	Blank		1	
4	Solvent Blank	Vial 2	D:\MassHunter\Data\Checkout\6460\SolventBlank002.d	Blank		1	
5	Solvent Blank	Vial 2	D:\MassHunter\Data\Checkout\6460\SolventBlank003.d	Blank		1	
6	Solvent Blank	Vial 2	D:\MassHunter\Data\Checkout\6460\SolventBlank004.d	Blank		1	
7	Solvent Blank	Vial 2	D:\MassHunter\Data\Checkout\6460\SolventBlank005.d	Blank		1	
8	Reserpine 500 fg	Vial 3	D:\MassHunter\Data\Checkout\6460\ReserpineCheckout001	Sample		1	
9	Reserpine 500 fg	Vial 3	D:\MassHunter\Data\Checkout\6460\ReserpineCheckout002	Sample		1	
10	☒ Reserpine 500 fg	Vial 3	D:\MassHunter\Data\Checkout\6460\ReserpineCheckout003	Sample		1	
11	☒ Reserpine 500 fg	Vial 3	D:\MassHunter\Data\Checkout\6460\ReserpineCheckout004	Sample		1	
12	☒ Reserpine 500 fg	Vial 3	D:\MassHunter\Data\Checkout\6460\ReserpineCheckout005	Sample		1	

Worklist

For Help, press F1 6460 ESI_AJT Pos MRM Checkout.m D:\MassHunter\...\6460 ESI_AJT Pos MRM Checkout.wkl is running Disk free space: 189.65 GB CAP NUM SCRL

Figure 22 Reserpine checkout worklist

- 1 Review the worklist to be sure that the method and data paths are correct, and that the data file names given in the worklist are unique and have not already been acquired.
- 2 Run the worklist.
- 3 When the worklist is finished, calculate signal-to-noise for each injection:
 - a Load each solvent blank and reserpine sample data file into the Qualitative Analysis program.
 - b Generate Extracted Ion Chromatograms of the 195.1 ion.





Figure 23 Extract Chromatogram dialog box

- c** Smooth the extracted chromatogram by setting **Smoothing Function** to **Quadratic/Cubic Savitzky-Golay**, and **Function width** to **11 points**.



- d** Integrate each reserpine peak, and click **Calculate Signal to Noise**.
- e** Calculate the signal-to-noise using **Height**.

3 Installation Completion and System Verification

Step 5. Verify the Triple Quad sensitivity

f *For sensitivity verification with a standard checkout:*

Under Noise Measurement, for **Noise definition**, select **RMS** and select **X 1**.

For **Specific noise regions**, type 0.100 – 0.350. (Make sure that the noise region does not include the reserpine peak.)




Figure 24 Calculate Signal-to-Noise dialog box (for standard checkout)

g *For sensitivity verification performed under G6061A:*

Under Noise Measurement, for **Noise definition**, select **Auto-RMS** (available in MassHunter Qualitative Analysis B.05.00 and higher).

Set the **Noise region boundary** to **Start time 0 minute** and **End time 1 minute**. (Make sure that the noise region does not include the reserpine peak.)

Set the **Noise region width** to **0.100 minute**.




Figure 25 Calculate Signal-to-Noise dialog box (for G6061A Performance Specification Service checkout)

- 4** Generate a printout of each signal-to-noise calculation report for each solvent blank and reserpine injection. Include the chromatogram in the printout.
- 5** Open the Excel spreadsheet **D:\MassHunter\Support\Checkout\Sensitivity Checkout Report.xls**. Fill in the values to calculate the average signal to noise and save the spreadsheet.

The average signal-to-noise must be greater than that listed in [Table 10](#).

3 Installation Completion and System Verification

Step 5. Verify the Triple Quad sensitivity

Table 10 Minimum average signal-to-noise (standard installation checkout)

Instrument Model	Source	Average Signal-to-Noise (1 x RMS)
6490	AJS ESI	10,000:1
6460	AJS ESI	1000:1
6430	ESI	300:1
6420	ESI	150:1
6420	Multimode	150:1

Table 11 Minimum average signal-to-noise for bid specification (sold as part of G6061A)

Instrument Model	Source	Average Signal-to-Noise (Auto-RMS)
6490	AJS ESI	50,000:1
6460	AJS ESI	10,000:1
6430	ESI	3,000:1
6420	ESI	1,500:1

6 Generate a printout of the Excel signal-to-noise report.

	A	B	C	D	E	F
1	6460 Triple Quad LC/MS Sensitivity Checkout Report					
2						
3						
4	Instrument Serial Number:	US83310889				
5						
6	Date:	28-Aug-2008				
7						
8						
9						
10		Signal-to-Noise				
11	Blank #1	0.27				
12	Blank #2	0.38				
13	Blank #3	0.42				
14	Blank #4	0.3				
15	Blank #5	0.41				
16	Blank Ave. S/N =	0				
17						
18			Area	Height		
19	Reserpine #1:	1207.2	924.00	404.00		
20	Reserpine #2:	1368.5	973.00	419.00		
21	Reserpine #3:	1583.7	959.00	412.00		
22	Reserpine #4:	1132.6	1044.00	448.00		
23	Reserpine #5:	2086	1091.00	471.00		
24	Reserpine Ave. S/N =	1476	998.20	430.80	AVE	
25			67.80	27.94	STD	
26			6.79%	6.49%	RSD	
27						
28	Signal-to-Noise =	1475:1				
29	(Reserpine Ave. - Blank Ave.)	PASS				
30						
31						

If verification in negative mode is required, verify the instrument with the parameters listed in [Table 12](#). Use the Agilent ES Negative Ion Performance Standard (p/n G1946-85005) that is included with your instrument.

Table 12 Electrospray Negative Mode

Sample	Acid Red 4 10 ng/μl in 50:50 water / IPA
Concentration after dilution	1 pg/μl
Injection volume	1 μl
Total sample amount injected	1 pg
Sample order number	G1946-85005
Solvents	50:50 isocratic 100% Water (no modifiers) 100% Acetonitrile (no modifiers)
Method	ESI Neg MRM Acid Red 4 Checkout.m
Worklist	ESI Neg MRM Acid Red 4 Checkout.wkl

To reset the Collision Cell gas flow rate

Most parameters like the Collision Cell gas flow rate should work with values set at the factory. In case the default flow rate does not give a high vacuum gauge reading in the range of 5 to 5.8×10^{-5} torr for a 6490, 1.9 to 2.3×10^{-5} for a 6460/6430 (two turbo pumps) and 2.7 to 3.3×10^{-5} for 6420/6410 (one turbo pump), you will need to do two steps:

- Determine the optimal flow rate (DAC steps) to achieve the correct high vacuum gauge reading.
- Enter this value into a Data Acquisition configuration file.

1 Determine the optimal Flow Rate:

- a Open a DOS window (click **Start** > **Run**, type `cmd` and click **OK**).
- b Type `telnet 192.168.254.12 5123` to open a “back door” into the Triple Quad's MS Interface Card firmware.
- c Press **Enter**.
- d Type `msepeek 1605` at the `>` prompt to view the current value (in DAC Steps) for the Flow Rate.

You should get a response like this:

```
[61577.0]: backdoor() command:msepeek 1605(12)
[61577.0]: MSE_READ: address = 1605, value = 520.
```

520 is the factory default value.

- e Find the high vacuum gauge reading in the **Cell** tab of the Manual Tune program.
- f Use the syntax below to adjust the value for register 1605 until the high vacuum gauge reading in the Manual Tune program is in the range of 1.9 to 2.3×10^{-5} :

```
>msepoke 1605, value
```

For example, to set the 1605 register value to 550, type the following and press **Enter**.

```
>msepoke 1605, 550
```

You may need to try numbers that are above and below the default value.

- g Write this number down.

- 2 Edit the acquisition parameter file so that the newly determined value will be applied whenever Data Acquisition starts up and will be used for tuning and data acquisition:

- a Open the **D:\MassHunter\instr\default** folder, and then open the appropriate .xml file with Notepad:

6490	G6490AHWParams.xml
6460	G6460AHWParams.xml
6430	G6430AHWParams.xml
6420	G6420AHWParams.xml
6410	G6410BHWParams.xml

- b Locate the set point for the Collision Cell gas flow between the `<flowrate>` and `</flowrate>` tags. In this example, the set point for the flow rate is 520:

```
<cellFlow>
  <flowA>1</flowA>
  <flowB>0</flowB>
  <flowrate>520</flowrate>
</cellFlow>
```

- c Change the flow rate set point to the value that you wrote down as the optimal flow rate. In this example, the value is changed to 550:

```
<cellFlow>
  <flowA>1</flowA>
  <flowB>0</flowB>
  <flowrate>550</flowrate>
</cellFlow>
```

- d Click **File > Save** to save the file.

- e Click **File > Exit** to close the file.

- 3 Restart the Data Acquisition program:

- a Close the Data Acquisition program.
- b Close all MassHunter engines from the MassHunter Acquisition Engine launcher. (Click the **Acq System Launcher** icon.)
- c Restart the MassHunter engines.
- d Start the Data Acquisition program.

3 Installation Completion and System Verification

To reset the Collision Cell gas flow rate

You should now read the high vacuum value which you targeted for and can proceed with the installation as described.


4 Verify the Collision Cell gas flow:

a In the DOS window, type the following and press **Enter**:

```
>msepeek 1605
```

b Close the DOS window:

```
>bye
```



4 Installation and Verification of Other Sources

G1947B APCI Source 82

To install the G1947B APCI source 82

To verify the operation of the G1947B APCI source 83

G1978B Multimode Source 84

To install the G1978B multimode source 84

To verify the operation of the G1978B multimode source 86

G1971B APPI Source 87

To install the G1971B APPI source 87

To verify the operation of the G1971B APPI source 88

HPLC-Chip/MS Cube Interface 89

Step 1. Prepare performance verification sample 89

Step 2. Prepare the HPLC-Chip/MS 91

Step 3. Analyze the results 94

This chapter contains the steps needed to install and verify the operation of the APCI, multimode and APPI sources.

If the APCI, multimode or APPI or HPLC-Chip/MS source was purchased with the Triple Quad instrument, you will need to install the source and verify its operation.

You can complete only a check tune with the different source types: G1948B and G1978B. Make sure you use the correct tune calibrant.



G1947B APCI Source

To install the G1947B APCI source

- 1 Install these parts from the APCI enablement kits that ship with the G1947B APCI source into the Aux module on the Triple Quad instrument:
 - APCI High Voltage Power Supply (p/n G1946-80058)
- 2 Pour the Electrospray calibrant back into its original bottle or another suitable container, rinse the calibrant bottle with acetonitrile, pour the APCI calibrant into the calibrant bottle, and attach the calibrant bottle back onto the CDS.
- 3 Remove the electrospray source and install the APCI source:
 - a Put the nebulizer into the nebulizer adjustment fixture that is supplied in the shipping kit. Check that the nebulizer needle is properly adjusted. Make sure that the nebulizer needle is even with the end of the nebulizer nozzle.
 - b Install the nebulizer in the spray chamber.

WARNING

Surfaces can be extremely hot.

- c Remove the previously installed spray chamber. Install the standard spray shield if not already installed.
- d Orient the standard spray shield so the off-axis hole is at the 12 o'clock position by loosening the two T10 screws that secure the end plate, orienting the spray shield, and re-tightening the two T10 screws (do not overtighten).
- e Install the APCI spray chamber on the spray chamber mount, close the spray chamber, and fasten the latch. If needed, adjust the latch to ensure that the O-ring seals completely. Use a 1/4-inch × 5/16-inch wrench to loosen the lock nut, adjust the latch to the proper fit, and then tighten the lock nut so that the latch maintains its adjustment.
- f Connect the APCI corona and vaporizer heater cables to the connector on the left side panel of the instrument.

- g Connect the 1/8-inch nebulizing gas tubing from the LC/MS mainframe to the nebulizer gas fitting.
- h Connect the PEEK tubing from the selection valve (inside the front cover) to the top of the nebulizer.

To verify the operation of the G1947B APCI source

There is no sensitivity checkout with the G1947B APCI source on the Triple Quad instrument. To verify its proper operation, run a manual tune with the G2432A APCI/APPI Calibration Solution.

- 1 Install the G2432A APCI/APPI Calibration Solution.
- 2 Check the tuning of the APCI source:
 - a In Data Acquisition program, set **Context** to **Tune**.
 - b Open the autotune file that was generated with the electrospray or multimode source.
 - c In the **Manual Tune** tab, mark the **Calibrant** check box, and then acquire the data.
 - d Verify that the masses in the **MS1** and **MS2** tabs are correct for the loaded source before you manually tune the peaks.
 - e Verify that you have sufficient abundance for the tune peaks, that the tune peaks have peak widths of 0.7 ± 0.05 amu, and that mass assignments are within ± 0.15 amu of the target mass assignments.

If needed, click **Adjust Gain & Offset** to establish the Mass Axis and Peak Width values. Repeat these for both MS1 and MS2 tabs.
- 3 Save the tune file.

G1978B Multimode Source

For more information on installing the multimode source, please refer to the multimode user guide and set-up guide.

To install the G1978B multimode source

If the multimode source is not already installed on your Triple Quad instrument, do these steps.

- 1 Install these parts from the multimode and APCI enablement kits that ship with the G1978B multimode source into the Aux module on the Triple Quad instrument:
 - APCI High Voltage Power Supply (p/n G1946-80058)
 - Multimode HV board cable (p/n G1960-60858)

CAUTION

Do not install the G1960-65015 Multimode HV board on the 6460 or you can damage the AJS ESI. The Multimode HV board is already installed on the 6460 and is backward-compatible with the G1978B multimode source.

The Multimode Power Data cable is already installed on the 6460.

- 2 Pour the electrospray tuning mix back into its original bottle or another suitable container, rinse the calibrant bottle with acetonitrile, pour the MMI-L tuning mix into the calibrant bottle, and attach the calibrant bottle back onto the CDS.
- 3 Remove the electrospray source and install the multimode source:
 - a Put the nebulizer into the nebulizer adjustment fixture that is supplied in the shipping kit. Check that the nebulizer needle is properly adjusted. Make sure that the nebulizer needle is even with the end of the nebulizer nozzle.
 - b Install the nebulizer in the spray chamber.

WARNING

Surfaces can be extremely hot.

- c** Remove the previously installed spray chamber. Install the multimode spray shield if not already installed.
- d** Orient the standard spray shield so the field-shaping electrode is at the 6 o'clock and 9 o'clock positions by loosening the two T10 screws that secure the end plate, orienting the spray shield, and re-tightening the two T10 screws (do not overtighten).
- e** Install the multimode spray chamber on the spray chamber mount, close the spray chamber, and fasten the latch. If needed, adjust the latch to ensure that the O-ring seals completely. Use a 1/4-inch × 5/16-inch wrench to loosen the lock nut, adjust the latch to the proper fit, and then tighten the lock nut so that the latch maintains its adjustment.
- f** Connect the multimode high voltage, APCI corona, and vaporizer heater cables to the connector on the left side panel of the instrument.
- g** Connect the 1/8-inch nebulizing gas tubing from the LC/MS mainframe to the nebulizer gas fitting.
- h** Connect the PEEK tubing from the selection valve (inside the front cover) to the top of the nebulizer.

To verify the operation of the G1978B multimode source

- 1 Install the G1969-85020 MMI-L Low Concentration Tuning Mix.
- 2 Check the tuning of the multimode source:
 - a In Data Acquisition program, set **Context** to **Tune**.
 - b Open the autotune file that was generated with the electrospray or multimode source.
 - c In the **Manual Tune** tab, mark the **Calibrant** check box, and then acquire the data.
 - d Verify that the masses in the **MS1** and **MS2** tabs are correct for the loaded source before you manually tune the peaks.
 - e Verify that you have sufficient abundance for the tune peaks, that the tune peaks have peak widths of 0.7 ± 0.05 amu, and that mass assignments are within ± 0.15 amu of the target mass assignments.

If needed, click **Adjust Gain & Offset** to establish the Mass Axis and Peak Width values. Repeat these for both MS1 and MS2 tabs.
- 3 Save the tune file.

G1971B APPI Source

To install the G1971B APPI source

- 1 Pour the Electrospray calibrant back into its original bottle or another suitable container, rinse the calibrant bottle with acetonitrile, pour the APCI/APPI calibrant into the calibrant bottle, and attach the calibrant bottle back onto the CDS.
- 2 Remove the currently installed source and install the APPI source:
 - a Put the nebulizer into the nebulizer adjustment fixture that is supplied in the shipping kit. Check that the nebulizer needle is properly adjusted. Make sure that the nebulizer needle is even with the end of the nebulizer nozzle.
 - b Install the nebulizer in the spray chamber.

WARNING

Surfaces can be extremely hot.

- c Remove the previously installed spray chamber. Install the standard spray shield if not already installed.
- d Orient the standard spray shield so the off-axis hole is at the 12 o'clock position by loosening the two T10 screws securing the end plate, orienting the spray shield, and re-tightening the two T10 screws (do not overtighten).
- e Install the APPI spray chamber on the spray chamber mount, close the spray chamber, and fasten the latch. If needed, adjust the latch to ensure that the O-ring seals completely. Use a ¼-inch × 5/16-inch wrench to loosen the lock nut, adjust the latch to the proper fit, and then tighten the lock nut so that the latch maintains its adjustment.
- f Connect the vaporizer heater cable to the connector on the left side panel of the instrument.
- g Connect the 1/8-inch nebulizing gas tubing from the LC/MS mainframe to the nebulizer gas fitting.

To verify the operation of the G1971B APPI source

To verify the operation of the G1971B APPI source

- Do *not* use the APPI USB to Serial Converter Cable (p/n 8121- 1013).

- 3** Connect the APPI power supply cable to the APPI DB9 power connector and screw it in. This connector contains both serial and power interfaces.
- 4** Connect the other end of the serial cable to the Serial connector on the smart card interface, which is located on the left side of the instrument chassis.
- 5** Plug the APPI power supply into an AC outlet using the power cord supplied with the APPI interface kit.

There is no sensitivity checkout with the G1971 APPI source on the Triple Quad instrument. To verify its proper operation, run a manual tune with the G2432A APCI/APPI Calibration Solution.

- 1 Install the G2432A APCI/APPI Calibration Solution.
- 2 Check the tuning of the APPI source:
 - a In Data Acquisition program, set **Context** to **Tune**.
 - b Open the autotune file that was generated with the electrospray or multimode source.
 - c In the **Manual Tune** tab, mark the **Calibrant** check box, and then acquire the data.
 - d Verify that the masses in the **MS1** and **MS2** tabs are correct for the loaded source before you manually tune the peak.
 - e Verify that you have sufficient abundance for the tune peaks, that the tune peaks have peak widths of 0.7 ± 0.05 amu, and that mass assignments are within ± 0.15 amu of the target mass assignments.

If needed, click **Adjust Gain & Offset** to establish the Mass Axis and Peak Width values. Repeat these for both MS1 and MS2 tabs.

- ### 3 Save the tune file.

HPLC-Chip/MS Cube Interface

Follow the steps in this section to do verification with the HPLC-Chip/MS Cube Interface.

Step 1. Prepare performance verification sample

- 1 Prepare the HSA peptides standard (p.n. G2455-85001) by adding 500 μL of 15:85 acetonitrile:water with 0.1% formic acid to the HSA peptide standard (500 pmol/vial).

- 2 Mix well on a vortex mixer to completely dissolve the standard.

The resulting stock solution is 1 pmol/ μL and contains 7 peptides. Only 1 peptide will be used for the quantitation checkout.

- 3 Create each of the dilutions listed in [Table 13](#):

- a In a conical bottom polypropylene autosampler vial, put the volume of standard in [Table 13](#).
- b Add the volume of 15:85 acetonitrile:water with 0.1% formic acid solvent in [Table 13](#).

The conical bottom polypropylene autosampler vials (p/n 5188-2788) and top with the appropriate caps (p/n 5182-0541) are included in the shipment.

Sample E is optional and is used for collecting information.

Table 13 Dilutions for HSA peptides standard

Sample	Volume of Standard	Volume of Solvent	Final Concentration
A	10 μL of 1 pmol/ μL solution	90 μL	100 fmol/ μL
B	10 μL of A	90 μL	10 fmol/ μL
C	10 μL of B	90 μL	1 fmol/ μL
D	10 μL of C	90 μL	100 amol/ μL
E	10 μL of D	90 μL	10 amol/ μL

4 Installation and Verification of Other Sources

Step 1. Prepare performance verification sample

- 4** Create a blank vial of 15:85 acetonitrile:water with 0.1% formic acid.

Use the same conical vials and caps used for the HSA samples.

- 5** Prepare a vial containing 60:40 acetonitrile:water with 1% TFA to clean the injector before doing low level samples.

This solution works well for solubilizing hydrophobic peptides. If a clean blank cannot be achieved after running several injections of this solution, remove the needle seat (with seat capillary) and place it seat-side down in a beaker with this solution, and then sonicate for 5 to 10 minutes.

Step 2. Prepare the HPLC-Chip/MS

- Copy the files from the Support folder to the Methods and Worklists folders according to [Table 14](#).

Table 14 Methods and worklists to copy

Copy...	From...	To...
6400 HPLC-Chip HSA Checkout.m	D:\MassHunter\Support\Checkout\6400 Methods and Worklist\Pos\HPLC-Chip HSA Checkout	D:\MassHunter\Methods
6400 HPLC-Chip HSA Checkout.wkl	D:\MassHunter\Support\Checkout\6460 Methods and Worklist\Pos\HPLC-Chip HSA Checkout	D:\MassHunter\Worklists

If you want to create your own method and worklist, do these steps:

- Create a worklist in MassHunter Acquisition program to run the prepared samples. See [Figure 26](#).

If you are time-limited, you can do a single injection of each, but triplicate injections are preferred. The 10 amol HSA level is optional.

Remember to condition a new chip with 500 fmol of HSA 3 times. Then clean up with blanks and the solubilizing solution.

	Sample Name	Sample Position	Method	Data File	Sample Type	Level Name	Inj Vol (µl)	Comment
1	blank	P1-D1	43mm-9min-hsa-mrm-575-unit.m	blank-r001.d	Blank		1	
2	blank	P1-D1	43mm-9min-hsa-mrm-575-unit.m	blank-r002.d	Blank		1	
3	blank	P1-D1	43mm-9min-hsa-mrm-575-unit.m	blank-r003.d	Blank		1	
4	100 amol/ul HSA	P1-F5	43mm-9min-hsa-mrm-575-unit.m	100amol-r001.d	Calibration	1	1	
5	100 amol/ul HSA	P1-F5	43mm-9min-hsa-mrm-575-unit.m	100amol-r002.d	Calibration	1	1	
6	100 amol/ul HSA	P1-F5	43mm-9min-hsa-mrm-575-unit.m	100amol-r003.d	Calibration	1	1	
7	1 fmol/ul HSA	P1-F6	43mm-9min-hsa-mrm-575-unit.m	1fmol-r001.d	Calibration	2	1	
8	1 fmol/ul HSA	P1-F6	43mm-9min-hsa-mrm-575-unit.m	1fmol-r002.d	Calibration	2	1	
9	1 fmol/ul HSA	P1-F6	43mm-9min-hsa-mrm-575-unit.m	1fmol-r003.d	Calibration	2	1	
10	10 fmol/ul HSA	P1-F7	43mm-9min-hsa-mrm-575-unit.m	10fmol-r001.d	Calibration	3	1	
11	10 fmol/ul HSA	P1-F7	43mm-9min-hsa-mrm-575-unit.m	10fmol-r002.d	Calibration	3	1	
12	10 fmol/ul HSA	P1-F7	43mm-9min-hsa-mrm-575-unit.m	10fmol-r003.d	Calibration	3	1	
13	100 fmol/ul HSA	P1-F8	43mm-9min-hsa-mrm-575-unit.m	100fmol-r001.d	Calibration	4	1	
14	100 fmol/ul HSA	P1-F8	43mm-9min-hsa-mrm-575-unit.m	100fmol-r002.d	Calibration	4	1	
15	100 fmol/ul HSA	P1-F8	43mm-9min-hsa-mrm-575-unit.m	100fmol-r003.d	Calibration	4	1	

Figure 26 HPLC-Chip/MS Worklist

4 Installation and Verification of Other Sources

Step 2. Prepare the HPLC-Chip/MS

2 Create the QQQ acquisition method as shown below:

Nanoflow Pump

Solvents	A1: 3:97 acetonitrile:water with 0.1% formic acid B1: 90:10 acetonitrile:water with 0.1% formic acid	
Flow	600 nL/min. with fast gradient solvent consumption	
Max pressure	175 bar	
Stop time	9 minutes	
Post time	0 minutes	
Gradient	Time	%B
	0	3.0
	7	70.0
	7.1	3

Capillary Pump

Solvent	A1: 97:3 water:acetonitrile with 0.1% formic acid
Flow	4.00 µL/min. (primary flow 200 µL/min.)
Max pressure	175 bar

Micro wellplate sampler

Injection volume	1 µL
Needle wash	enabled for 5 sec in flushport
Needle flush solvent	20:80 methanol:water with 0.1% formic acid
Bottom sensing	On
Vial offset	0
ALS Therm	On, 40°C

HPLC-Chip MS Interface

Injection flush volume	4 µL
Pumps	Intelligent Sample Loading is activated.
Chip Cube Timetable	Valve to enrichment at 7.5 min.

QQQ MRM Method

Ionization mode	positive HPLC-Chip
Time filter	on
Time Filter Width	0.07
Dry gas flow	5 L/min.

Dry gas temp325 °C

Capillary Voltage1700 to 1800 V is typical

Time segments1

Delta EMV0 V (can go up to 200 V if more noise is desired)

MRM scan segments:

Acquisition	Source	Chromatogram	Instrument	Diagnostics							
Scan segments											
Compound Name	ISTD?	Precursor Ion	MS1 Res	Product Ion	MS2 Res	Dwell	Fragmentor	Collision Energy	Cell Accelerator Voltage	Polarity	
Peptide 3	☐	575.5	Wide	937.5	Unit	100	380	20	5	Positive	
Peptide 3	☐	575.5	Wide	823.4	Unit	100	380	20	5	Positive	
Peptide 3	☐	575.5	Wide	694.4	Unit	100	380	20	5	Positive	

Step 3. Analyze the results

- 1 In the Quantitative Analysis program, create a batch and add the samples.
- 2 Create a calibration curve and use the Curve Fit Assistant to optimize the curve to obtain a best fit that includes all points and optimizes Accuracy for the data points.




Figure 27 HSA Calibration Curve Example

You can use the Qualitative Analysis program to calculate the signal-to-noise ratio.

- 3 For 100 amol/μL in the 575.50 to 937.50 transition range, check that the minimum signal-to-noise ratio (defined from 0.5 to 3.0 minutes) is 100:1 (3x RMS noise).




Figure 28 Calculated Signal-to-noise for 575.5 m/z -> ** transitions and a TIC MRM.

4 Installation and Verification of Other Sources

Step 3. Analyze the results





Figure 29 Mass Spectrum for MRM transition 575.5 m/z -> **



5 Other Setup Tasks

- To start the Data Acquisition program 98
- To stop MassHunter Engines 98
- To stop MassHunter processes in case of unrecoverable errors 98
- To view the log book 99
- To configure the network interface card 100
- To remove the MassHunter Workstation software 101
- To reconfigure the instrument 102

This chapter contains tasks that are related to setting up your Triple Quad instrument.

For complete information on using the your Triple Quad instrument, refer to the list of resources found in the *Quick Start Guide*.



To start the Data Acquisition program

You can start the Data Acquisition program in one of two ways:

- Double-click the **Data Acquisition** icon in the Agilent MassHunter program group, or
- Click **Start > Programs > Agilent > MassHunter Workstation > Data Acquisition**.

CAUTION

Do not start Data Acquisition as administrator. (*Do not* right-click **Data Acquisition** and select **Run as administrator**.)

To stop MassHunter Engines

- 1 Close the **Data Acquisition** program.
- 2 Right-click the **Engine Launcher** icon in the system tray.
- 3 Click **Shutdown Engines** to stop the engines.

To start the MassHunter Engines, restart the Data Acquisition program. Alternatively, to start the MassHunter Engines without starting the Data Acquisition program, click **Start Engines**.

To stop MassHunter processes in case of unrecoverable errors

When you stop MassHunter processes, loss of data can occur.

- 1 Click **Programs > Agilent > MassHunter Workstation > Acq Tools > Remove MassHunter Processes**.
- 2 Click **Yes** when you are asked whether you want to remove all MassHunter processes.

Only the Data Acquisition processes are stopped.

To view the log book

- Double-click the **Acq System Logbook** icon in the **Agilent MassHunter** program group, or click **Start > Programs > Agilent > MassHunter Workstation > Acq System Logbook**.

You can also start the Logbook Viewer program within the MassHunter Data Acquisition program.

To configure the network interface card

The computer that is shipped with the Triple Quad instrument is equipped with two network cards (LAN-cards). The first LAN-card is for connection to the “house network” (via 10/100/1000 Base-T network card). The IP address for this card can either be “static” or dynamically assigned (e.g. obtained by a DHCP server). By default, TCP/IP is configured as the network protocol for this card, which is built-in to the PC.

The second LAN-card (in slot #1) is dedicated for instrument communication with the Triple Quad instrument and the Agilent 1100/1200 Series LC system. *This network card should NOT be connected to the site LAN.*

To verify if the Instrument LAN network card settings are correct, follow these steps.

- 1 Click **Start > Settings > Control Panel**.
- 2 Double-click **Network Connections**.
- 3 Right-click the icon for the network card for the house LAN, click **Properties**, and then configure the network card as needed.

Because this network card is independent from the card used for instrument control, it may be configured using a static IP address or using a DHCP server.

- 4 Right-click the icon for the network card for the Triple Quad instrument, click **Properties**, and then check that these parameters are set:
 - Subnet Mask 255.255.255.0
 - Gateway - leave blank

NOTE

Because network configurations vary widely from site to site, get assistance from an IT representative when you connect to the site network. Configuration by an Agilent Customer Engineer is on a best-effort basis only.

To remove the MassHunter Workstation software

- 1 If you have any data, methods, libraries, or other files or directories that you want to save, then move or copy them to a folder outside of those to be deleted.
- 2 Close all MassHunter Workstation programs, including the Data Acquisition, Qualitative Analysis and Quantitative Analysis programs.
- 3 Click **Start > Control Panel > Programs and Features**.
- 4 Remove these programs:
 - **Agilent MassHunter WorkStation Data Acquisition**
 - **Agilent MassHunter WorkStation Qualitative Analysis**
 - **Agilent MassHunter WorkStation Quantitative Analysis**
 - **Agilent MassHunter WorkStation Quantitative Analysis Reporting**
- 5 Close *all* applications.
- 6 Reboot the computer. You are now ready to install new software.

To reconfigure the instrument

If you have more than one stack of HPLC modules from which to choose, or a new supported LC module is installed, the Instrument Configuration registry must be cleared for MassHunter Workstation.

- 1 Put the Triple Quad instrument into standby mode.
- 2 Close the MassHunter Acquisition program.
- 3 Stop MassHunter processes.

See “To stop MassHunter processes in case of unrecoverable errors” on page 98.

- 4 Start the Instrument Configuration program (**Programs > Agilent > MassHunter Workstation > Acq Tools > Instrument Configuration**).
- 5 Mark the **Mass Spectrometer** check box, and then click the Triple Quad instrument. Leave all other check boxes cleared.




Figure 30 Instrument Configuration

- 6 Click **OK**.
- 7 Start the MassHunter Workstation Acquisition program and confirm that only the MS instrument is available.

- 8** Close the Data Acquisition program.
- 9** Stop MassHunter processes.
- 10** Open Instrument Configuration again.
- 11** Click the Triple Quad instrument and mark the **Agilent LC System Access** check box.




Figure 31 Instrument Configuration dialog box

- 12** Click **OK**.


The registry has now been cleared and configured for any new LC module(s).

Additional configuration steps may be required to set up LC devices.

elemental compositions and calculated masses of sodium trifluoroacetate (TFA) ion clusters that can result from electrospray ionization under positive-ion ($[(CF_3COONa)_nNa]1$) and negative-ion ($[(CF_3COONa)_nCF_3COO]2$) modes.

5 Other Setup Tasks

To reconfigure the instrument



6 Reference

6460/6430 Reserpine Checkout Method Acquisition Parameters	106
6490 Reserpine Checkout Method Acquisition Parameters	110

This chapter contains acquisition parameters for the reserpine checkout method.



6460/6430 Reserpine Checkout Method Acquisition Parameters

The following graphics give the default acquisition parameters for the 6460 and 6430 POS MRM Reserpine Checkout method.




Figure 32 Properties tab

6460/6430 Reserpine Checkout Method Acquisition Parameters

Method Editor

6460 ES1_AJT Pos MRM Checkout.m Apply

Properties DA **Sampler** Sampler Pretreatment Binary Pump Column Comp. VWD QQQ

Sampler (G1329B)

Injection Mode

Injection volume: 2.00 μ L

☒ Standard injection
☐ Injection with needle wash

Needle wash

Location:

Stop time **Post time**

☒ As Pump/No Limit ☒ Off
☐ 1.00 min ☐ 1.00 min

Advanced

Auxiliary

Draw speed: 200 μ L/min
Eject speed: 200 μ L/min
Draw position: 0.0 mm

High throughput

☐ Enable Optimization
☒ Prefetch Vial
☐ Overlap Injection Cycle
0.00 minutes after Injector

WorkList Method Editor **Sample Run**

Figure 33 HiP Sampler tab

Method Editor

6460 ES1_AJT Pos MRM Checkout.m Apply

Properties DA Sampler Sampler Pretreatment **Binary Pump** Column Comp. VWD QQQ

Binary Pump (G1312B)

Flow

0.400 mL/min

Solvents

A: 30.0 %
1 H₂O with 5 mM Formic Acid
2

B: 70.0 %
1 ACN
2

Pressure Limits

Min: 0.00 bar Max: 400.00 bar

Stop time **Post time**

☐ As Injector/No Limit ☒ Off
☒ 1.00 min ☐ 1.00 min

Advanced

Timetable (empty)

☐ function centric view

Time [min]	A [%]	B [%]	Flow [mL/min]	Max. Pressure Limit [bar]
0.00	30.0	70.0	0.400	400.00

Add Remove Clear All Clear Empty
Cut Copy Paste Shift Times min

WorkList Method Editor **Sample Run**

Figure 34 Binary Pump tab

6460/6430 Reserpine Checkout Method Acquisition Parameters




Figure 35 Column Comp. tab




Figure 36 QQQ > Acquisition tab (6460 only)

6460/6430 Reserpine Checkout Method Acquisition Parameters

Method Editor

6460 ESI_AJT Pos MRM Checkout.m Apply

Properties DA Sampler Sampler Pretreatment Binary Pump Column Comp. VWD **QQQ**

Tune file: atunes.TUNE.XML Browse...

Ion source: AJS ESI

Stop time: ☒ No limit/As Pump ☐ 2.0 min

Time filtering: ☒ Peak width 0.03 min

Time segments:

#	Start Time	Scan Type	Div Valve	Delta EMV (+)	Delta EMV (-)	Stored
1	0	MRM	To MS	250	0	☒

4.91 cycles/s 203.5 ms/cycle

Acquisition **Source** Chromatogram Instrument Diagnostics

Source parameters:

Gas Temp: 325 °C 319.875 °C Copy Paste

Gas Flow: 10 l/min 8.0 l/min

Nebulizer: 20 psi 27.0 psi

Sheath Gas Temp: 400 °C 380 °C

Sheath Gas Flow: 12 l/min 12.0 l/min

Capillary: Positive 4000 V Negative 3500 V 9997.555 nA

Nozzle Voltage: 0 V 0 V

Chamber Current: 0.52 µA

Paste to All Segments

WorkList Method Editor Sample Run

Figure 37 QQQ > Source tab for AJS ESI (6460 only)

Method Editor

6460 ESI_AJT Pos MRM Checkout.m Apply

Properties DA Sampler Sampler Pretreatment Binary Pump Column Comp. VWD **QQQ**

Tune file: atunes.TUNE.XML Browse...

Ion source: AJS ESI

Stop time: ☒ No limit/As Pump ☐ 2.0 min

Time filtering: ☒ Peak width 0.03 min

Time segments:

#	Start Time	Scan Type	Div Valve	Delta EMV (+)	Delta EMV (-)	Stored
1	0	MRM	To MS	250	0	☒

4.91 cycles/s 203.5 ms/cycle

Acquisition Source **Chromatogram** Instrument Diagnostics

Chromatograms:

Chromatogram Type	Label	Precursor Ion	Product Ion	Offset	Y-Range
MRM	609>195	609.3	195.1	10	5000
TIC	TIC			0	1E+07

WorkList Method Editor Sample Run

Figure 38 QQQ > Chromatogram tab

6490 Reserpine Checkout Method Acquisition Parameters

The following graphics give the default acquisition parameters for the 6490 POS MRM Reserpine Checkout method.




Figure 39 Properties tab




Figure 40 HiP Sampler tab

6490 Reserpine Checkout Method Acquisition Parameters

Properties DA HiP Sampler HiP Sampler Pretreatment **Binary Pump** Column Comp. DAD QQQ

Binary Pump (G4220A)

Flow
0.400 mL/min

Solvents

A: 1 100.0 % Water V.03 0.1% formic a
2 100.0 % Water V.03

B: 1 100.0 % Acetonitrile V.03 ACN
2 100.0 % Acetonitrile V.03

Pressure Limits
Min: 0.00 bar Max: 1,200.00 bar

Stoptime
☐ As Injector/No Limit
☒ 2.00 min

Posttime
☐ Off
☒ 1.00 min

Advanced
Timetable (4/100 events)

Time [min]	A [%]	B [%]	Flow [mL/min]	Max. Pressure Limit [bar]
0.00	90.00	10.00	0.400	1200.00
0.20	90.00	10.00	---	---
0.80	0.00	100.00	---	---
1.00	0.00	100.00	---	---
1.01	90.00	10.00	---	---

☐ function centric view

Add Remove Clear All Clear Empty
Cut Copy Paste Shift Times min

WorkList Method Editor Sample Run

Figure 41 Binary Pump tab

Properties DA HiP Sampler HiP Sampler Pretreatment **Binary Pump** **Column Comp. (G1316C)** DAD QQQ

Temperature

Left: ☐ Not Controlled ☒ 40.0 °C ☐ As Detector Cell

Right: ☐ Not Controlled ☒ 20.0 °C ☐ As Detector Cell ☐ Combined

Stoptime
☒ As Pump/Injector
☐ 1.00 min

Posttime
☐ Off
☒ 1.00 min

Advanced
Enable Analysis

☒ when front door open

Left: ☐ With any temperature ☒ When temperature is within ± 0.8 °C

Right: ☐ With any temperature ☒ When temperature is within ± 0.8 °C

Timetable (empty)

Figure 42 Column Comp. tab

6 Reference
6490 Reserpine Checkout Method Acquisition Parameters

Method Editor

6490_ESI_AJT_Pos_MRM_Checkout.m Apply

Properties DA HIP Sampler HIP Sampler Pretreatment Binary Pump Column Comp. DAD **QQQ**

Acquisition Source Chromatogram Instrument Diagnostics

Scan segments

Compound Group	Compound Name	ISTD?	Precursor Ion	MS1 Res	Product Ion	MS2 Res	Dwell	Fragmentor	Collision Energy	Cell Accelerator Voltage	Polarity
	Reserpine	☐	609.3	Wide	195.1	Unit	200	380	41		Positive

Tune file: atunes.TUNE.XML
Browse ...

Ion source: AJS ESI

Stop time: ☐ No limit/As Pump ☐ 1 min

Time filtering: ☒ Peak width 0.03 min

Time segments

#	Start Time	Scan Type	Div Valve	Delta EMV (+)	Delta EMV (-)	Stored
1	0	MRM	To MS	200	0	☒

4.91 cycles/s 203.5 ms/cycle

WorkList Method Editor Sample Run

Figure 43 QQQ > Acquisition tab

Method Editor

6490_ESI_AJT_Pos_MRM_Checkout.m Apply

Properties DA HIP Sampler HIP Sampler Pretreatment Binary Pump Column Comp. DAD **QQQ**

Acquisition **Source** Chromatogram Instrument Diagnostics

Source parameters

Gas Temp:	200 °C	225 °C
Gas Flow:	12 l/min	14.9 l/min
Nebulizer:	20 psi	15.0 psi
Sheath Gas Temp:	400 °C	125 °C
Sheath Gas Flow:	12 l/min	3.0 l/min

Positive Negative

Capillary:	3500 V	3000 V	210 nA
Nozzle Voltage:	0 V	1500 V	
Chamber Current:			0.17 µA

iFunnel parameters

	Positive	Negative
High Pressure RF	150 V	90 V
Low Pressure RF	60 V	60 V

Copy Paste

Paste to All Segments

Tune file: atunes.TUNE.XML
Browse ...

Ion source: AJS ESI

Stop time: ☐ No limit/As Pump ☐ 1 min

Time filtering: ☒ Peak width 0.03 min

Time segments

#	Start Time	Scan Type	Div Valve	Delta EMV (+)	Delta EMV (-)	Stored
1	0	MRM	To MS	200	0	☒

4.91 cycles/s 203.5 ms/cycle

Figure 44 QQQ > Source tab for AJS ESI

6490 Reserpine Checkout Method Acquisition Parameters

6460 ESL_AJT_Pos MRM Checkout.m

Properties DA HiP Sampler HiP Sampler Pretreatment Binary Pump Column Comp. DAD QQQ

Tune file: atunes.TUNE.XML [Browse ...]

Stop time: ☐ No limit/As Pump ☐ 2.0 min

Ion source: ESI ☒ Agilent Jet Stream

Time filtering: ☒ Peak width 0.03 min

Time segments:

#	Start Time	Scan Type	Div Valve	Delta EMV (+)	Delta EMV (-)	Stored
▶ 1	0	MRM	To MS	250	0	☒

4.91 cycles/s 203.5 ms/cycle

Acquisition Source **Chromatogram** Instrument Diagnostics

Chromatograms

Chromatogram Type	Label	Precursor Ion	Product Ion	Offset	Y-Range
▶ MRM	609>195	609.3	195.1	10	5000
TIC	TIC			0	1E+07

Figure 45 QQQ > Chromatogram tab

www.agilent.com

In This Book

This book contains
installation, configuration,
verification, and start-up
tasks to operate your
Triple Quad LC/MS
System.

© Agilent Technologies, Inc. 2012

Revision A, November 2012



G3335-90150



Agilent Technologies